


# Certified Reference Material

## Reference material certificate

### Hexadecane

TraceCERT®  
Traceable Certified Reference Materials

<b>Product no.:</b>	CRM78293
<b>Lot no.:</b>	BCCM5846
<b>Description of CRM:</b>	Liquid neat material
<b>Expiry date:</b>	JAN 2029
<b>Storage:</b>	15-25°C; storage under Argon
<b>Chemical formula:</b>	$\text{C}_{16}\text{H}_{34}$
<b>Molecular mass:</b>	226.44 g/mol
<b>CAS No.:</b>	544-76-3



Sample	<b>Certified value <math>\pm</math> Expanded uncertainty, <math>U_{CRM}=k \cdot u_{CRM}</math> (<math>k=3.2</math>)<sup>[1][2]</sup> as mass fraction</b>
--------	---

<b>Hexadecane</b>	<b>0.996 g/g <math>\pm</math> 0.010 g/g</b> (99.6 % $\pm$ 1.0 %)
-------------------	--

<b>Metrological traceability:</b>	Traceable to the SI through an unbroken chain of comparisons via NIST PS1 (Benzoic acid) <sup>[3]</sup> Details see "Certification process details" on page 2.
<b>Measurement method:</b>	The certified value is established by high-resolution quantitative NMR measurements (qNMR) in accordance with ISO/IEC 17025. <sup>[4]</sup>
<b>Intended use:</b>	Use this certified reference material (CRM) as calibrant for chromatography or any other analytical technique.
<b>Minimum sample size:</b>	There is no recommended minimum sample size. The substance is liquid at room-temperature and therefore homogeneous under these conditions. Potential inhomogeneity is covered by the expanded measurement uncertainty.
<b>Instructions for handling and correct use:</b>	The CRM should be stored in the original bottle. Warm to room temperature before opening. Shake gently before use, to ensure complete homogeneity. After use the bottle should be tightly closed and protected from excessive moisture and light.
<b>Accreditation:</b>	Sigma-Aldrich Production GmbH is accredited by the Swiss Accreditation Service SAS as reference material producer under no. SRMS 0001 in accordance with international standard ISO 17034. <sup>[5]</sup>
<b>Certificate issue date:</b>	27 JAN 2025



ISO 17034  
SRMS 0001

Dr. P. Zell – Approving Officer  
(Quality Assurance)



<b>Health and safety information:</b>	Please refer to the Safety Data Sheet for detailed information about the nature of any hazard and appropriate precautions to be taken.
<b>Packaging:</b>	Brown glass bottle
<b>Starting material details:</b>	Starting materials are checked with different analytical techniques (e.g. 1D/2D NMR methods) for suitability in terms of purity and compatibility with solvent and qNMR reference.


#### Certification process details:

In order to guarantee highest reliability of this **TraceCERT** CRM a multi-component approach was applied whereby the certified value is determined by high-resolution quantitative NMR measurements (qNMR on a 600 MHz NMR spectrometer).<sup>[4][6]</sup> The whole workflow follows the ISO 17034 guidelines. The certificate is designed in accordance with ISO 33401.<sup>[7]</sup>

The certified values are confirmed by extended analytical data. These data are not covered by the scope of accreditation but determined following best practices in analytical measurements.

The high precision weighing steps are performed in accordance with ISO/IEC 17025<sup>[4]</sup> using ultra-micro balances certified by DKD and calibrated with OIML Class E2 weights.

Absolute content determination by qNMR is performed using 4-5 separate samples of the candidate substance which are each spiked with an adequate amount of internal reference and then immediately dissolved in deuterated solvent. In most cases 16-32 scans are recorded for every sample with a  $^1\text{H}$  relaxation time of  $d1=60$  seconds. Quantification of the candidate content is directly calculated from the  $^1\text{H}$ -NMR peak areas and the initial mass of the candidate and reference substance. After ANOVA the resulting standard deviation is included into the uncertainty calculation of the certified value. Extensive stability and homogeneity tests are considered for certification.<sup>[8]</sup>




**Homogeneity assessment:** Homogeneity of the material is tested by qNMR measurements using 4-5 subsamples which are taken from different positions in the entire bulk material. The recommended minimal sample size is taken for all the homogeneity test samples. Analysis of variance (ANOVA) results are included into the calculation of content uncertainty of this CRM.

**Stability assessment:** An accelerated stability test is performed with samples which are stored above the recommended storage temperature. The material is tested by qNMR double determination at appropriate time intervals, e.g. 3 months. The long-term stability test is performed with samples which are stored at the recommended storage temperature and applying qNMR double determination at appropriate time intervals, e.g. 24 months.

## Uncertainty evaluation:

The uncertainty contributions are illustrated by the following cause-effect diagram.




The combined standard uncertainty is calculated by combination of the standard uncertainties of the input estimates according to Eurachem/CITAC Guide "Quantifying Uncertainty in Analytical Measurement" and ISO 17034.<sup>[2][5]</sup>

Expanded uncertainty is then calculated to a confidence level of 95 %, typically by multiplying with a coverage factor  $k=3.2$ .

## Qualitative NMR analysis:

Identity assessment is carried out by NMR analysis in accordance with ISO/IEC 17025.<sup>[4]</sup>



## References:

- [1] ISO Guide 35:2017, "Reference materials - Guidance for characterization and assessment of homogeneity and stability"
- [2] Eurachem/CITAC Guide, 3<sup>rd</sup> Ed. (2012), "Quantifying uncertainty in analytical measurement"
- [3] Eurachem/CITAC Guide, 2<sup>nd</sup> Ed. (2019), "Metrological traceability in chemical measurement"
- [4] The accredited testing laboratory STS 0490 performs the measurements and weighing steps for the certification of this CRM under ISO/IEC 17025:2017, "General requirements for the competence of testing and calibration laboratories"
- [5] ISO 17034:2016, "General requirements for the competence of reference material producers"
- [6] Weber M, Hellriegel C, Rueck A, Sauermoser R, Wuethrich J, Accred. Qual. Assur. 18 (2013) 91-98
- [7] ISO 33401:2024, "Reference materials - Contents of certificates, labels and accompanying documentation"
- [8] Weber M, Hellriegel C, Rueck A, Wuethrich J, Jenks P, JPBA 93 (2014) 102-110

## Certificate of analysis revision history:

Certificate version	Date	Reason for version
01	27 JAN 2025	Initial version

The most recent version of the Certificate is available online ([www.sigmaaldrich.com](http://www.sigmaaldrich.com))

## Disclaimer:

The purchaser must determine the suitability of this product for its particular use. Sigma-Aldrich Production GmbH makes no warranty of any kind, express or implied, other than its products meet all quality control standards set by Sigma-Aldrich Production GmbH. We do not guarantee that the product can be used for a special application.

The vibrant M, Supelco, TraceCERT and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. All other trademarks are the property of their respective owners. Detailed information on trademarks is available via publicly accessible resources. © 2018 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the US and Canada.

