

QUALITY CONTROL RANGES
Milliplex Rat Liver Injury Magnetic Panel
Kit Catalog# RL11MAG-92K
Control Catalog # RL11-6092 Lot# RL192-104 and RL192-204



Analyte	QC Level	Expected Range	Units	Analyte	QC Level	Expected Range	Units
ARG1	Control 1	0.95 – 2.1	ng/mL	5'-NT	Control 1	0.30 – 0.62	ng/mL
	Control 2	10.2 – 21	ng/mL		Control 2	3.8 – 7.8	ng/mL
GOT1	Control 1	0.66 – 1.4	ng/mL	SDH	Control 1	11 – 22	ng/mL
	Control 2	7.9 – 16	ng/mL		Control 2	59 – 122	ng/mL
GSTa	Control 1	0.16 – 0.34	ng/mL				
	Control 2	2.0 – 4.2	ng/mL				

RL11-6092: RL192-104.204 (05 January 2023)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) www.emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA