

QUALITY CONTROL RANGES
Canine Kidney Toxicity Expanded Magnetic Bead Panel 1
Kit Catalog# CKT1MAG-97K
Control Catalog# CKT1-6097-1
Lot# CKT1-108 and CKT1-208



<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
KIM-1	Control 1	4.5 – 9.3	pg/mL
	Control 2	37 – 78	pg/mL
Clusterin	Control 1	6.3 – 13	ng/mL
	Control 2	54 – 112	ng/mL
Osteopontin	Control 1	0.12 – 0.25	ng/mL
	Control 2	0.55 – 1.1	ng/mL
IL-8	Control 1	0.33 – 0.68	ng/mL
	Control 2	2.8 – 5.7	ng/mL
Cystatin C	Control 1	0.06 – 0.12	ng/mL
	Control 2	0.51 – 1.1	ng/mL
Lipocalin-2/NGAL	Control 1	0.22 – 0.45	ng/mL
	Control 2	1.1 – 2.2	ng/mL
MCP-1	Control 1	0.52 – 1.1	ng/mL
	Control 2	2.4 – 5.3	ng/mL

CKT1-6097-1: CKT1-108.208 (24 February 2023)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA