# Sigma-Aldrich

Technical Bulletin

# Starch Assay Kit

Catalogue number MAK522

# **Product Description**

Starch, chemical formula  $(C_6H_{10}O_5)_n$ , is a polysaccharide carbohydrate consisting of a large number of glucose units joined together by glycosidic bonds. All plant seeds and tubers contain starch present in the form of amylose and amylopectin. Starch is the most consumed polysaccharide in the human diet. Some starches are digested very quickly and cause a rapid and large rise in blood sugar. Others are digested more slowly, and some starch, called resistant starch, is not digested in the small intestine at all, and thus causes little or no blood sugar rise. Simple, direct and automation-ready procedures for measuring starch concentrations find wide applications in research and drug discovery.

The Starch Assay Kit uses a single Working Reagent that combines the enzymatic breakdown of starch and the detection of glucose in one step. The color intensity of the reaction product at 570 nm or fluorescence intensity at  $\lambda$ ex =530 nm /  $\lambda$ em = 585 nm is directly proportional to the starch concentration in the sample. This simple convenient assay is carried out at room temperature and takes only 30 min.

The linear detection range of the kit is 2 to 200  $\mu g/mL$  starch for colorimetric assays and 0.2 to 20  $\mu g/mL$  for fluorometric assays.

# Components

The kit is sufficient for 100 colorimetric or fluorometric assays in 96-well plates.

•	Assay Buffer Catalogue Number MAK522A	12 mL
•	Dye Reagent Catalogue Number MAK522B	120 µL
•	Enzyme A Catalogue Number MAK522C	1 vial
•	Enzyme B Catalogue Number MAK522D	120 μL
•	Standard (50 mg/mL) Catalogue Number MAK522E	50 μL

# Equipment Required but Not Provided

- Pipetting devices and accessories (For example multichannel pipettor)
- Multiwell plate reader
- Clear flat-bottom 96-well plates for colorimetric assay or black flat-bottom 96-well plates for fluorometric assay. Cell culture or tissue culture treated plates are not recommended.
- Centrifuge tubes Dounce tissue grinder set (Catalog Number D9063 or equivalent)

# Precautions and Disclaimer

For R&D use only. Not for drug, household, or other uses. Please consult the Safety Data Sheet for information regarding hazards and safe handling practices.

# Storage/Stability

The kit is shipped on wet ice. Store components at -20 °C.



# Preparation Instructions

Briefly centrifuge small vials prior to opening. Equilibrate all components to room temperature prior to use.

Enzyme A: Reconstitute Enzyme A by adding 120  $\mu$ L Assay Buffer to the Enzyme A tube. Make sure Enzyme A is fully dissolved by pipetting up and down. During the experiment, keep thawed enzymes in a refrigerator or on ice. Store reconstituted Enzyme A at -20 °C and use within 1 month.

Note: 1. This assay is based on a kinetic reaction, the use of a multi-channel pipettor for adding the working reagent is recommended.

2. SH-group containing reagents (For example - DTT,  $\beta$ -mercaptoethanol) may interfere with this assay and should be avoided in sample preparation.

#### Procedure

All Samples and Standards should be run in duplicate.

#### Reagent Preparation

#### Soluble Starch:

- Use 5-10 mg of the Sample and grind followed by washing off any free glucose and small oligosaccharides with 1 mL 90% ethanol.
- 2. Warm to 60 °C for 5 min with occasional vortexing. Centrifuge at 10,000g for 2 min.
- 3. Decant the supernatant.
- 4. Repeat the wash twice. Remove ethanol.

#### Soluble starch in the pellet form:

- 1. Extract with 1 mL  $H_2O$  incubated in a boiling water bath for 5 min.
- 2. Spin 10,000 x g for 2 min.
- 3. The supernatant is soluble starch and resistant starch remains as an insoluble pellet.

#### Resistant Starch:

- 1. After extracting soluble starch, extract the water insoluble pellet with 0.2 mL DMSO and heat in boiling water bath for 5 min.
- 2. Dilute sample 1:100 in H<sub>2</sub>O prior to assay.

### Colorimetric Procedure

Transfer 10  $\mu L$  of Sample into separate wells of a clear flat-bottom microplate.

### Standard curve preparation:

Prepare a 200  $\mu$ g/mL Standard by diluting 5  $\mu$ L of 50 mg/mL Standard with 1.245 mL purified water. Further, dilute Standards as mentioned in Table 1.

Table 1. Dilution of Standard

Well No.	200 μg/mL STD (μL)	Purified Water (µL)	Starch Concentration (µg/mL)
1	200	0	200
2	150	50	150
3	100	100	100
4	50	150	50
5	0	200	0

Transfer 10  $\mu L$  of the Standards and Samples into separate wells of a clear flat-bottom microplate.

Note: If the Sample contains glucose, transfer an additional 10  $\mu$ L Sample to another well for the Sample Blank.

## Working Reagent Preparation

For each Standard and Sample well, prepare enough Working Reagent according to Table 2.

Table 2.

Preparation of Working Reagents

Reagent	Working Reagent	Blank Working Reagent
Assay Buffer	90 μL	85 µL
Enzyme A	1 μL	-
Enzyme B	1 μL	1 μL
Dye Reagent	1 µL	1 µL

Transfer 90  $\mu L$  Working Reagent into each reaction well. Add 90  $\mu L$  Blank Working Reagent to the Sample Blank Wells. Tap plate to mix.

#### Measurement

- 1. Incubate for 30 mins at Room Temperature.
- 2. Read the optical density at 570nm (550-585 nm)

# Fluorometric Procedure

1. Prepare a 200  $\mu$ g/mL Standard by diluting 5  $\mu$ L of 50 mg/mL Standard with 1.245 mL purified water. Further, dilute Standards as mentioned in Table 1.

Table 1. Dilution of Standard

Well No.	200 μg/mL STD (μL)	Purified Water (µL)	Starch Concentration (µg/mL)
1	20	180	20
2	15	185	15
3	10	190	10
4	5	195	5
5	0	200	0

- 2. Repeat the same procedure as performed in the Colorimetric procedure except use a use a black flat-bottom microplate.
- 3. Incubate for 30 min at room temperature and read fluorescence at

 $\lambda_{ex}$  = 530 nm /  $\lambda_{em}$  = 585 nm.

#### Results

- 1. Subtract the blank value from all the standard values.
- 2. Plot the  $\Delta F$  or  $\Delta OD$  against the standard concentrations.
- Determine the slope and calculate Starch concentration of samples using the below given equation:


Starch ( $\mu$ g/mL) =  $\frac{R_{Sample} - R_{Blank}}{Slope}$ 

#### Where:


 $R_{Sample}$  = Optical Density (OD) or Fluorescence Intensity (F) of the Sample

 $R_{Blank}$  = Optical Density (OD) or Florescence Intensity (F) of the Blank

**Figure 1.** Typical standard curve from Colorimetric Assay.



**Figure 2.**Typical standard curve from Fluorometric Assay.



#### **Notice**

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

The information in this document is subject to change without notice and should not be construed as a commitment by the manufacturing or selling entity, or an affiliate. We assume no responsibility for any errors that may appear in this document.

#### **Technical Assistance**

Visit the tech service page at SigmaAldrich.com/techservice.

#### Terms and Conditions of Sale

Warranty, use restrictions, and other conditions of sale may be found at SigmaAldrich.com/terms.

#### Contact Information

For the location of the office nearest you, go to SigmaAldrich.com/offices.

The life science business of Merck operates as MilliporeSigma in the U.S. and Canada.

