



Quality Control Ranges

MILLIPLEX[®] MAP Human Sepsis

Magnetic Bead Panel 3

Kit Catalog # HSP3MAG-63K

Control Cat # HSP3-6063-3 Lot# HSP3-110.210

<u>Analyte</u>	<u>QC Level</u>	<u>Lot #</u>	<u>Expected Range</u>	<u>Units</u>
HELA2-MAG	Control I	HSP3-110	224 - 464	pg/mL
	Control II	HSP3-210	1120 - 2327	pg/mL
HLTF-MAG	Control I	HSP3-110	2286 - 4748	pg/mL
	Control II	HSP3-210	10468- 21740	pg/mL
HNGAL-MAG	Control I	HSP3-110	325 - 676	pg/mL
	Control II	HSP3-210	1595 - 3312	pg/mL
HRES-MAG	Control I	HSP3-110	114 - 237	pg/mL
	Control II	HSP3-210	581 - 1208	pg/mL
HTSP1-MAG	Control I	HSP3-110	2106 - 4373	pg/mL
	Control II	HSP3-210	10902 - 22643	pg/mL

HSP3-6063-3: HSP3-110.210 (6 Feb 2025)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) www.emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA