


SIGMA QUALITY CONTROL TEST PROCEDURE**Enzymatic Assay of GLUTAMINASE¹**
(EC 3.5.1.2)
(From E. coli)**PRINCIPLE:****CONDITIONS:** T = 37°C, pH = 4.9, A_{340nm}, Light path = 1 cm**METHOD:** Spectrophotometric Stop Rate Determination**REAGENTS:**

- A. 100 mM Sodium Acetate Buffer, pH 4.9 at 37°C
(Prepare 100 ml in deionized water using Sodium Acetate, Trihydrate, Sigma Prod. No. S-8625. Adjust to pH 4.9 at 25°C with 1 M HCl.)
- B. 5 mM Sodium Acetate Buffer, pH 6.0 at 37°C
(Prepare 10 ml in deionized water using Sodium Acetate, Trihydrate, Sigma Prod. No. S-8625. Adjust to pH 6.0 at 25°C with 1 M HCl.)
- C. 80 mM L-Glutamine Solution
(Prepare 10 ml in Reagent A using L-Glutamine, Sigma Prod. No. G-3126. **PREPARE FRESH.**)
- D. Glutaminase Solution
(Immediately before use, prepare a solution containing 5 units/ml of Glutaminase in cold Reagent B.)
- E. Ammonia Diagnostic Kit (171-20)
(Use Ammonia Reagent, Sigma Stock No. 171-20.)
- F. Ammonia Diagnostic Kit (170-4)
(Use L-Glutamate Dehydrogenase, Sigma Stock No. 170-4)

Enzymatic Assay of GLUTAMINASE¹
(EC 3.5.1.2)
(From E. coli)

PROCEDURE:

Step 1:

Pipette (in milliliters) the following reagents into suitable tubes:

	<u>Test</u>	<u>Blank</u>
Reagent A (Sodium Acetate Buffer)	0.4	0.4
Reagent C (L-Glutamine)	0.5	0.5

Equilibrate to 37°C. Then add:

Reagent D (Glutaminase)	0.1	-----
Reagent B (Sodium Acetate Buffer)	-----	0.1

Immediately mix by swirling and incubate at 37°C for 15 minutes. Then add:

Cold Deionized Water	9.0	9.0
----------------------	-----	-----

Immediately use as Test and Blank solutions in Step 2.

Step 2:

Pipette (in milliliters) the following reagents into suitable cuvettes:

	<u>Test</u>	<u>Blank</u>
Reagent E (Ammonia Reagent)	1.0	1.0

Equilibrate to 25°C. Then add:

Test Solution (10 ml)	0.1	-----
Blank Solution (10 ml)	-----	0.1

Monitor the $A_{340\text{nm}}$ until constant, using a suitably thermostatted spectrophotometer. Record the initial $A_{340\text{nm}}$ for both Test and Blank. Then add:

Reagent F (L-Glutamate Dehydrogenase)	0.01	0.01
---------------------------------------	------	------

Immediately mix by inversion and record the decrease in $A_{340\text{nm}}$ until completion (approximately 5 minutes). Obtain the final $A_{340\text{nm}}$ for both the Test and Blank.

Enzymatic Assay of GLUTAMINASE¹
(EC 3.5.1.2)
(From E. coli)

CALCULATIONS:

$$\Delta A_{340\text{nm}} \text{ Test} = A_{340\text{nm}} \text{ Test Final} - A_{340\text{nm}} \text{ Test Initial}$$

$$\Delta A_{340\text{nm}} \text{ Blank} = A_{340\text{nm}} \text{ Blank Final} - A_{340\text{nm}} \text{ Blank Initial}$$

$$\text{Units/ml enzyme} = \frac{\Delta A_{340\text{nm}} \text{ Test} - \Delta A_{340\text{nm}} \text{ Blank} (10 \text{ ml})(1.11)}{(6.22) (15) (0.1) (0.1 \text{ ml})}$$

6.22 = Millimolar extinction coefficient of β -NADH at 340nm

15 = Reaction time of Step 1

RM = Reaction Mixture of Step 1

10 = Final volume of Step 1

0.1 = Volume from Step 1 used in Step 2

1.11 = Step 2 Reaction Mix volume

0.1 = Volume (in milliliters) of enzyme used

$$\text{Units/mg solid} = \frac{\text{units/ml enzyme}}{\text{mg solid/ml enzyme}}$$

$$\text{Units/mg protein} = \frac{\text{units/ml enzyme}}{\text{mg protein/ml enzyme}}$$

UNIT DEFINITION:

One unit will deaminate 1.0 μ mole of L-glutamine per minute at 37°C at pH 4.9.

FINAL ASSAY CONCENTRATION:

In a 1.0 ml reaction mix, the final concentrations are 90 mM sodium acetate, 40 mM L-glutamine and 0.5 units glutaminase.

NOTES:

1. This assay is not for the use with glutaminase from porcine kidney.
2. All products and stock numbers, unless otherwise indicated, are Sigma product and stock numbers.

Sigma warrants that the above procedure information is currently utilized at Sigma and that Sigma products conform to the information in Sigma publications. Purchaser must determine the suitability of the information and products for its particular use. Upon purchase of Sigma products, see reverse side of invoice or packing slip for additional terms and conditions of sale.