



Quality Control Ranges  
**MILLIPLEX® MAP Mouse Myokine Magnetic Bead Panel**  
**Catalog # MMYOMAG-74K**  
**Control Catalog # MMY-6074 Lot# MMY-103.203**

Hormone	QC Level	Expected Range	Units	Hormone	QC Level	Expected Range	Units
RBDNF	Control 1	7.1 – 15	pg/mL	MSTN	Control 1	5917 – 12290	pg/mL
	Control 2	76 – 158	pg/mL		Control 2	61257 – 127225	pg/mL
EPO	Control 1	288 – 598	pg/mL	IRISN	Control 1	501 – 1040	pg/mL
	Control 2	3048 – 6331	pg/mL		Control 2	5165 – 10728	pg/mL
IL-6	Control 1	124 – 257	pg/mL	LIF	Control 1	34 – 71	pg/mL
	Control 2	1276 – 2650	pg/mL		Control 2	330 – 686	pg/mL
IL-15	Control 1	861 – 1789	pg/mL	OSM	Control 1	220 – 457	pg/mL
	Control 2	8323 – 17287	pg/mL		Control 2	2074 – 4307	pg/mL
FGF21	Control 1	61 – 127	pg/mL	OSTCRN	Control 1	31 – 64	pg/mL
	Control 2	669 – 1390	pg/mL		Control 2	314 – 653	pg/mL
FLSTN	Control 1	2317 – 4812	pg/mL	SPARC	Control 1	1754 – 3643	pg/mL
	Control 2	22905 – 47573	pg/mL		Control 2	18290 – 37987	pg/mL

MMY-6074: MMY-103.203 (30 March 2021)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • [emdmillipore.com](http://emdmillipore.com)

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.  
© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.