

**Human Circulating Cancer Biomarker
MILLIPLEX® MAP Magnetic Bead Panel 3
Kit Catalog # HCCBP3MAG-58K
Control Catalog # HCC3-6058-3
Lot# HCC3-108 and HCC3-208**



<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
Galectin	Control 1	0.015 - 0.031	ng/mL
	Control 2	0.40 - 0.84	ng/mL
Cathepsin D	Control 1	1.2 - 2.5	ng/mL
	Control 2	24 - 51	ng/mL
FAP	Control 1	0.31 - 0.64	ng/mL
	Control 2	6.3 - 13	ng/mL
MIA	Control 1	0.45 - 0.9	ng/mL
	Control 2	9.8 - 20	ng/mL
MPO	Control 1	0.12 - 0.25	ng/mL
	Control 2	2.5 - 5.2	ng/mL
SHBG	Control 1	0.14 - 0.29	nM
	Control 2	2.9 - 6.0	nM
IGFBP3	Control 1	0.4 - 0.8	ng/mL
	Control 2	15 - 32	ng/mL
Ferritin	Control 1	0.14 - 0.29	ng/mL
	Control 2	2.9 - 6.1	ng/mL

HCC3-6058-3: HCC3-108.208 (07 August 2020)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.