

Product Information Sheet

Histopaque®-1077

Catalog Number 10771

Product Description

Histopaque®-1077 is a sterile, endotoxin tested solution of polysucrose and sodium diatrizoate, adjusted to a density of 1.077 g/mL. This ready-to-use medium facilitates rapid recovery of viable lymphocytes and other mononuclear cells from small volumes of whole blood. Histopaque-1077 is suitable for human lymphocyte antigen (HLA) typing¹ and as the initial isolation step prior to enumeration of T, B, and 'null' lymphocytes.² It may also be employed in the preparation of pure lymphocyte suspensions for cell culture and cytotoxicity assays.³

Histopaque-1077 is a sterile solution of polysucrose, 57 g/L, and sodium diatrizoate, 90 g/L.

Density: 1.076-1.078 g/mL

Endotoxin: \leq 3 EU/cm³

pH: 8.8-9.0

Precautions and Disclaimer

For R&D use only. Not for drug, household, or other uses. Please consult the Safety Data Sheet for information regarding hazards and safe handling practices.

Storage/Stability

Store the product at 2–8 °C. Histopaque-1077 has an expiration period of 3 years. Reagent label bears expiration date.

Preparation Instructions

Specimen Collection - Collect blood in preservative-free anticoagulant (EDTA or heparin) or use defibrinated blood. For best results, blood should be processed within 2 hours.

Procedure

Anticoagulated blood is layered onto Histopaque-1077. During centrifugation, erythrocytes are aggregated by polysucrose and rapidly sediment. Granulocytes become slightly hypertonic, which increases their sedimentation rate, resulting in pelleting at the bottom of the centrifuge tube. Lymphocytes and other mononuclear cells remain at the plasma/ Histopaque-1077 interface. Erythrocyte contamination is negligible. Most extraneous platelets are removed by low speed centrifugation during the washing steps.

- To a 15 mL conical centrifuge tube, add 3 mL of Histopaque-1077 and bring to room temperature.
- 2. Carefully layer 3 mL of whole blood onto the Histopaque-1077.
- Centrifuge at 400 × g for exactly 30 minutes at room temperature.
 Centrifugation at lower temperatures, such as 4 °C, may result in cell clumping and poor recovery.

<u>Note:</u> Make sure brake and acceleration are on lowest setting on centrifuge, harsh braking and acceleration may affect layer separation.



- After centrifugation, carefully aspirate the upper layer with a Pasteur pipette to within 0.5 cm of the opaque interface containing mononuclear cells. Discard upper layer.
- 5. Carefully transfer the opaque interface with a Pasteur pipette into a clean conical centrifuge tube.
- Wash the cells by adding 10 mL of isotonic phosphate buffered saline solution or appropriate cell culture medium and mix by gently drawing in and out of a Pasteur pipette.
- 7. Centrifuge at $250 \times g$ for 10 minutes.
- 8. Aspirate the supernatant and discard.
- Resuspend cell pellet with 5 mL of isotonic phosphate buffered saline solution or appropriate cell culture medium and mix by gently drawing in and out of a Pasteur pipette.
- 10. Centrifuge at 250 \times g for 10 minutes.
- Repeat steps 8, 9, and 10, discard the supernatant and resuspend the cell pellet in 0.5 mL of isotonic phosphate buffered saline solution or appropriate cell culture medium.

Results

Erythrocytes and granulocytes should pellet to the bottom of the centrifuge tube.

Mononuclear cells should band at the interface between the Histopaque-1077 and the plasma.

If observed results vary from expected results, please contact MilliporeSigma Technical Service for assistance.

References

- Amos, D.B., and Pool, P., "HLA typing" in Manual of Clinical Immunology, Rose, N.R., and Friedman, H., eds., American Society for Microbiology, (Washington, DC: 1976) pp. 797-804.
- Winchester, R.J., and Ross, G., "Methods for enumerating lymphocyte populations" in Manual of Clinical Immunology, Rose, N.R., and Friedman, H., eds., American Society for Microbiology, (Washington, DC: 1976) pp. 64-76.
- 3. Thorsby, E., and Bratlie, A., "A rapid method for preparation of pure lymphocyte suspensions."
 Histocompatibility Testing, Terasaki, P.I., ed., 665-666 (1970).



Notice

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

The information in this document is subject to change without notice and should not be construed as a commitment by the manufacturing or selling entity, or an affiliate. We assume no responsibility for any errors that may appear in this document.

Contact Information

For the location of the office nearest you, go to SigmaAldrich.com/offices.


Technical Service

Visit the tech service page on our web site at SigmaAldrich.com/techservice.

Standard Warranty

The applicable warranty for the products listed in this publication may be found at SigmaAldrich.com/terms.

10771 Product Information Sheet Rev 12/2021



The life science business of Merck KGaA, Darmstadt, Germany

operates as MilliporeSigma in the U.S. and Canada.

