


Technical Bulletin

Peroxidase Assay Kit

Catalogue number **MAK506**

Product Description

Peroxidases catalyze the following oxidation-reduction reactions:



For many peroxidases the optimal substrate is hydrogen peroxide (H_2O_2), but others are more active with organic hydroperoxides such as lipid peroxides. In the cell, peroxidases destroy toxic hydroxide radicals that are formed as byproducts during aerobic respiration. The peroxidases represent a large family of enzymes that are found in animals (for example myeloperoxidase-like enzymes), plants, fungi and bacteria (cytochrome-c peroxidase like enzymes such as horseradish peroxidase).

Simple, direct and automation-ready procedures for determining peroxidase activity find wide applications. The Peroxidase Assay Kit uses H_2O_2 and an electron donor dye that forms resorufin during the peroxidase reaction. The optical density (570 nm) or fluorescence intensity at $\lambda_{\text{Ex}} = 530 \text{ nm} / \lambda_{\text{Em}} = 585 \text{ nm}$ is a direct measure of the enzyme activity.

The linear detection range of the kit is 2 to 50 U/L peroxidase for colorimetric assays and 0.1 to 5 U/L for fluorometric assays. The kit is suitable for Peroxidase activity determination in biological samples such as plasma, serum, urine, tissue, and culture media.

Components

The kit is sufficient for 100 colorimetric/fluorometric assays in 96-well plates.

• Assay Buffer Catalogue Number MAK506A	20 mL
• Stop Reagent Catalogue Number MAK506B	12 mL
• Resorufin Catalogue Number MAK506C	1.5 mL
• Dye Reagent Catalogue Number MAK506D	60 μL
• 3% Stabilized H_2O_2 Catalogue Number MAK506E	100 μL

Reagents and Equipment Required but Not Provided

- Pipetting devices and accessories (for example, multichannel pipettor)
- Multiwell plate reader.
- Clear flat-bottom 96-well plates for colorimetric assay or black flat-bottom 96-well plates for fluorometric assay. Cell culture or tissue culture treated plates are not recommended.
- 1.5 mL microcentrifuge tubes

Precautions and Disclaimer

For R&D use only. Not for drug, household, or other uses. Please consult the Safety Data Sheet for information regarding hazards and safe handling practices.

Storage/Stability

The kit is shipped on wet ice. Store components at $-20 \text{ }^{\circ}\text{C}$.

Preparation Instructions

Briefly centrifuge small vials prior to opening. Equilibrate all components to room temperature prior to use.

H₂O₂ Reagent: Dilute 3% H₂O₂ in Assay Buffer to 0.6% and use within one hour. Use freshly diluted 0.6% H₂O₂.

Note: This assay is based on a kinetic reaction, the use of a multi-channel pipettor for adding the Working Reagent and Stop Reagent is recommended.

Procedure

All Samples and Standards should be run in duplicate.

It is prudent to test multiple sample dilutions to ensure activity is in the linear range.

Working Reagent Preparation

Mix enough reagents for the number of assays to be performed. For each well, prepare 96 μ L of Working Reagent according to Table 1.

Table 1.

Preparation of Working Reagent

Reagent	Volume
Assay Buffer	95 μ L
Dye reagent	0.5 μ L
H ₂ O ₂ (0.6%)	0.5 μ L

Colorimetric Assay

1. Add 100 μ L purified water and 100 μ L Resorufin into two wells of a clear flat bottom 96-well plate.
2. Transfer 10 μ L purified water (sample blank), 10 μ L sample to separate wells.
3. Add 90 μ L Working Reagent into each well. Tap plate to mix.
4. Incubate the plate for 10 minutes at room temperature.
5. Add 100 μ L Stop Reagent to all wells. Read the optical density at 570nm.

Fluorometric Procedure

1. Dilute the Resorufin 1:10 in purified water.
2. Resorufin into two wells of a black flat-bottom 96-well plate.
3. Transfer 10 μ L purified water and 10 μ L sample to separate wells.
4. Add 90 μ L Working Reagent into each well. Tap plate to mix.
5. Incubate the plate for 10 minutes at room temperature.
6. Add 100 μ L Stop Reagent to all wells. Tap plate to mix. Read fluorescence ($\lambda_{ex}/\lambda_{em}$ = 530/585 nm).

Note: If Sample OD or fluorescence values are higher than that of the Resorufin, dilute sample in Assay Buffer, repeat assay and multiply results by the dilution factor, n.

Results

The peroxidase activity in a sample is computed as follows:

Peroxidase activity

$$= \frac{R_{\text{Sample}} - R_{\text{Blank}}}{R_{\text{Resorufin}} - R_{\text{H}_2\text{O}}} \times \frac{[\text{Resorufin}] (\mu\text{M})}{t (\text{min})} \times \frac{\text{Reaction vol} (\mu\text{L})}{\text{Sample vol} (\mu\text{L})} \times n$$

$$= \frac{R_{\text{Sample}} - R_{\text{Blank}}}{R_{\text{Resorufin}} - R_{\text{H}_2\text{O}}} \times [\text{Resorufin}] (\mu\text{M}) \times n \text{ (U/L)}$$

Where:

R_{Sample} , R_{Blank} , $R_{\text{Resorufin}}$ and $R_{\text{H}_2\text{O}}$ are optical density or fluorescence readings of the Sample, Sample Blank, Resorufin and Water respectively.

n is the dilution factor.


The [Resorufin] is 50 μM for colorimetric assays and 5 μM for fluorometric assays.

The Reaction Vol is 100 μL and the Sample Vol is 10 μL .

Unit Definition: one unit of enzyme will catalyze the formation of 1 μmole resorufin per minute under the assay conditions.

Figure 1.

Example of Resorufin standard curves by colorimetric and fluorometric assay.



Notice

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

The information in this document is subject to change without notice and should not be construed as a commitment by the manufacturing or selling entity, or an affiliate. We assume no responsibility for any errors that may appear in this document.

Technical Assistance

Visit the tech service page at SigmaAldrich.com/techservice.

Terms and Conditions of Sale

Warranty, use restrictions, and other conditions of sale may be found at SigmaAldrich.com/terms.

Contact Information

For the location of the office nearest you, go to SigmaAldrich.com/offices.

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

MilliporeSigma, and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. All other trademarks are the property of their respective owners. Detailed information on trademarks is available via publicly accessible resources.
© 2022 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

mak506pis Rev 12/23

The logo for MilliporeSigma, featuring the word "MILLIPORE" in a bold, red, sans-serif font above the word "SIGMA" in a similar bold, red, sans-serif font.