

QUALITY CONTROL RANGES
MILLIPLEX[®] MAP Human Kidney Injury Magnetic Bead Panel 2
Kit Catalog # HKI2MAG-99K
Control Catalog # HKI2-6099-2
Lot # HKI2-110 and HKI2-210



<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
EGF	Control 1	0.06 - 0.13	ng/mL
	Control 2	0.63 - 1.3	ng/mL
Lipocalin-2/NGAL	Control 1	0.13 - 0.27	ng/mL
	Control 2	1.2 - 2.6	ng/mL
Albumin	Control 1	23 - 49	ng/mL
	Control 2	245 - 509	ng/mL
Clusterin	Control 1	32 - 66	ng/mL
	Control 2	304 - 631	ng/mL
Cystatin C	Control 1	0.07 - 0.15	ng/mL
	Control 2	1.0 - 2.0	ng/mL
OPN	Control 1	4.2 - 8.7	ng/mL
	Control 2	39 - 81	ng/mL
α1-Microglobulin	Control 1	6.1 - 13	ng/mL
	Control 2	63 - 132	ng/mL

Note: The Quality Control Ranges are generated using assay buffer provided in the kit.
Quality Control Ranges in culture media are not tested.

HKI2-6099-2: HKI2-110.210 (12 July 2023)
EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) www.emdmillipore.com
The M logo is a trademark of Merck KGaA, Darmstadt, Germany.
© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA