

**PPARG Hu-Cy3  
SmartFlare™  
RNA Detection Probe**



**Cat. # SF-1903**

**pack size: 50µL (250 rxns)**

Store at 2-8°C,  
after reconstitution store at 23-27°C  
**DO NOT FREEZE**

FOR RESEARCH USE ONLY  
NOT FOR USE IN DIAGNOSTIC PROCEDURES  
NOT FOR HUMAN OR ANIMAL CONSUMPTION

**Product Data Sheet**

page 1

**Accession #NM\_015869.4**

**Species Hu**

**Gene Aliases: GLM1; CIMT1;  
NR1C3; PPARG1; PPARG2;  
PPARgamma**

**Confirmation of PPARG SmartFlare Performance:**

PPARG SmartFlare probe has been tested in a buffer system to detect the release of the fluorophore in the presence of a complementary base pair sequence for each lot to confirm target specificity.

PPARG SmartFlare probe has also been tested in a cell model system and demonstrated increased fluorescence in cells expressing the target compared to a scrambled negative control SmartFlare (Figure 1).

**Storage and Handling:**

Material has been 0.22µm filtered. Stable for 5 years at 2-8°C degrees in lyophilized format ONLY. Room temperature is required for reconstituted product.

**Warning-after reconstitution product is sensitive to cold and hot temperatures, a stable room temperature of 23-27°C is recommended.**

**Handling Recommendations:**

Reconstitute with sterile nuclease free water in a drop wise fashion tap tube repeatedly to fully dissolve lyophilized material.

Upon reconstitution, store at room temperature for up to 1 year protected from light. Product must be handled with gloves as product can be absorbed through the skin.

**Mean Fluorescence Intensity (MFI) Values**

Unflared	Scrambled	PPARG
7.940000000000004	105.68000000000001	229.56

**Figure 1:** PPARG Mean Fluorescence Intensity (green) by flow cytometry in living THP-1 cells demonstrated a significant increase over unflared cells (blue) as well as scramble control (red). Data shown is representative.

**Recommended Cell Testing Protocol:**

**(example: 30,000 cells in a 200µL media volume within each well of a 96 well plate)**

- Reconstitute reagent in 50µL of sterile nuclease free water.
- Create a working solution based on your experiment by diluting 1:20 in sterile PBS
- Add 4µL directly to cells (at approx 80% confluency)
- Allow to incubate overnight for 16 hrs
- Detect using fluorescence detection platform of choice

**SPECIES LEGEND:** Hu Human Ms Mouse Rt Rat Rb Rabbit WR Most Common Vertebrates ( ) Predicted Reactivity

**Please visit [www.millipore.com/SmartFlare](http://www.millipore.com/SmartFlare) for additional product information.**

EMD Millipore Corporation, 28820 Single Oak Drive, Temecula, CA 92590, USA 1-800-437-7500

Technical Support: T: 1-800-MILLIPORE (1-800-645-5476) • F: 1-800-437-7502

**FOR RESEARCH USE ONLY.** Not for use in diagnostic procedures. Not for human or animal consumption. Purchase of this Product does not include any right to resell or transfer, either as a stand-alone product or as a component of another product. Any use of this Product for purposes other than research is strictly prohibited.

EMD Millipore, the M mark and SmartFlare are registered trademarks of Merck KGaA, Darmstadt, Germany.

© 2012 EMD Millipore Corporation, Billerica MA USA. All rights reserved.



We Buy 100% Certified Renewable Energy

