

Benchmark Scientific Inc. 116 Corporate Blvd. South Plainfield, NJ 07080 USA

## **DECLARATION OF CONFORMITY**

Benchmark Scientific, Inc. hereby declares that the device(s) described below are in conformity with the EC Directives listed here-in as of the date of this declaration. In the event of unauthorized modification of any of the devices listed below, this declaration becomes invalid.

**Device Name:** H1000-M (-E), H1010 (-E) and H1012 (-E)

**Device Model Number(s):** Incu-Shaker Mini, Incu-Shaker 10L and Incu-Shaker 10LR Shaking

Incubators

**Relevant EC Directives:** Low Voltage Directive 2006/95/EC

Electromagnetic compatibility (EMC) Directive 2004/108/EC

**Harmonized Standards:** EN 61010-1: 2001

EN 61010-2-010:2003 EN 61010-2-051:2003 EN 613236-1:2006

EN 61000-3-2:2006+A1:2009+A2:2009

EN 61000-3-3:2008

Manufacturer's Name: Benchmark Scientific, Inc.

PO Box 709 Edison, NJ 08818

**Date:** March 06, 2013

Marketing Manager

Phone: (908) 769-5555 Fax: (908) 222-1864 Email: info@BenchmarkScientific.com