



Quality Control Ranges
Human Neuroscience Magnetic Bead Panel 2
Kit Cat# HNS2MAG-95K

Control Cat# HNS2-6095-2
Lot# HNS2-104 and HNS2-204

<u>Analyte</u>	<u>Control Level</u>	<u>Expected Range</u>	<u>Units</u>
TREM2	Control 1	343 - 711	pg/mL
	Control 2	1832- 3805	pg/mL
ApoE4	Control 1	6493- 13485	pg/mL
	Control 2	36479 - 75764	pg/mL
Angiogenin	Control 1	61 - 127	pg/mL
	Control 2	340- 706	pg/mL
FABP3	Control 1	609 - 1265	pg/mL
	Control 2	3247 - 6744	pg/mL
Neurogranin	Control 1	108 - 225	pg/mL
	Control 2	642 - 1334	pg/mL
Ferritin	Control 1	160 - 333	pg/mL
	Control 2	799 - 1660	pg/mL

HNS2-6095-2: HNS2-104.204 (22 March 2021)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.