



## QUALITY CONTROL RANGES

### Human Leptin ELISA (Normal & Sensitive) Kit Catalog #: EZHL-80SK

#### Control Catalog #: E6081-K

Kit	Level	Lot #	Expected Range	
Human Leptin	QC 1	EXHL 114	3.1 – 6.5	ng/mL
	QC 2	EXHL 214	10 – 21	ng/mL
Sensitive Human Leptin	QC 1	EXHL 114	1.0 – 2.2	ng/mL
	QC 2	EXHL 214	3.4 – 7.1	ng/mL

E6081-K: EXHL-114.214 (07 February 2024)

EMD MILLIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) • [www.emdmillipore.com](http://www.emdmillipore.com)

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA