



**QUALITY CONTROL RANGES**  
**MILLIPLEX® MAP Rat Myokine Magnetic Bead Panel**  
**Catalog # RMYMAG-88K**  
**Control Catalog # RMY-6088 Lot# RMY-103.203**

Hormone	QC Level	Expected Range	Units	Hormone	QC Level	Expected Range	Units
BDNF	Control 1	82 - 171	pg/mL	FSTL-1	Control 1	1578 - 3277	pg/mL
	Control 2	732 - 1521	pg/mL		Control 2	13461 - 27958	pg/mL
EPO	Control 1	352 - 731	pg/mL	MSTN/GDF8	Control 1	17 - 35	ng/mL
	Control 2	2990 - 6211	pg/mL		Control 2	138 - 286	ng/mL
IL-15	Control 1	7.7 - 16	ng/mL	IRISN	Control 1	6.9 - 14	ng/mL
	Control 2	65 - 135	ng/mL		Control 2	59 - 123	ng/mL
FGF21	Control 1	57 - 119	pg/mL	LIF	Control 1	573 - 1189	pg/mL
	Control 2	535 - 1111	pg/mL		Control 2	4983 - 10350	pg/mL
Fractalkine	Control 1	68 - 141	pg/mL	Osterocrin	Control 1	118 - 246	pg/mL
	Control 2	533 - 1106	pg/mL		Control 2	867 - 1801	pg/mL
IL-6	Control 1	4.2 - 8.6	ng/mL	SPARC	Control 1	3.2 - 6.6	ng/mL
	Control 2	34 - 70	ng/mL		Control 2	29 - 59	ng/mL

**Note: The Quality Control Ranges are generated with overnight assay format using serum matrix provided in the kit.  
Quality Control values in culture media are not tested.**

RYM-6088: RMY-103.203 (09 September 9, 2022)

• EMD MILIPORE • 1-800-645-5476 • 1-800-645-5439 (FAX) • [www.emdmillipore.com](http://www.emdmillipore.com)

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA