3050 Spruce Street, St. Louis, MO 63103 USA
Tel: (800) 521-8956 (314) 771-5765 Fax: (800) 325-5052 (314) 771-5757
email: techservice@sial.com sigma-aldrich.com

Product Information

(-)-Cotinine solution 1.0 mg/mL±5% in methanol drug standard

Catalog Number **C0430** Lot Number SLBK4808V Storage Temperature 2–8 °C

CAS RN 486-56-6

Synonyms: (S)-1-Methyl-5-(3-pyridyl)-2-pyrrolidinone, S(–)-1-Methyl-5-(3-pyridyl)-2-pyrrolidinone

Product Description

Molecular Formula: C₁₀H₁₂N₂O Molecular Weight: 176.22

DEA Class: Exempt preparation of a non-controlled

substance

Concentration: 0.97 mg of (–)-cotinine/mL of GC grade methanol, concentration verified by UV.

Expiration Date: July 2017

Precautions and Disclaimer


This product is for R&D use only, not for drug, household, or other uses. Please consult the Material Safety Data Sheet for information regarding hazards and safe handling practices.

Storage/Stability

Store the product at $2-8\,^{\circ}\text{C}$. If stored at $2-8\,^{\circ}\text{C}$ in an airtight container, protected from light, decomposition of the product is less than 1% in 12 months. After opening, concentration may change due to loss of solvent.

CMW, MAM 01/16-1


70 eV Electron Ionization Mass Spectrum



Hewlett-Packard 6890 Series II Gas Chromatograph Hewlett-Packard 5973 Mass Selective Detector

Hewlett-Packard G1701DA MS ChemStation (HP-DOS series)

FT-IR



Nicolet 380 FT-IR Smart Orbit (ATR)

Ultraviolet Spectrum

Peak 261.0 nm

Absorbance 0.7406

Solvent GC grade methanol


Dilution 1:24 (v/v)


Capillary Gas Chromatography


Column Supelco SPB-1 (Catalog Number 24162) I.D. 0.20 mm Length 15 m Film Thickness 0.20 μ m Oven Temperature 50 °C (2 minutes), then 15 °C/minute to 300 °C, hold 11 minutes Injector Temperature 250 °C, Split

Detector Hewlett-Packard 5973

Mass Selective Detector (70 eV, 40-500 m/z)







Hewlett-Packard 6890