

Technical Bulletin

L-Amino Acid Assay Kit

Catalogue number MAK490

Product Description

L-Amino acids are essential molecules defined by the nature of their sidechain. There are 20 amino acids that are commonly used by most organisms to produce proteins and other biologically relevant molecules. While amino acids can occur in either the D- or L-enantiomer, only the L-form is used by cells.

The L-Amino Acid Assay Kit uses an enzyme-catalyzed oxidation of L-amino acids to convert a dye into a colored and fluorescent form. The absorbance at 570 nm or fluorescence intensity at $\lambda_{\text{Ex}}=530 \text{ nm}/\lambda_{\text{Em}}=585 \text{ nm}$ is directly proportional to the L-amino acid concentration in the sample.

The linear detection range of the kit is 3.3 – 500 μM for the colorimetric assay and 0.13 – 50 μM for the fluorometric assay. The kit is suitable for L-amino acid determination in serum, culture media, tissue homogenates, cell lysates, urine, food samples, etc.

Components

The kit is sufficient for 100 colorimetric or fluorometric assays in 96-well plates.

• Assay Buffer Catalogue Number MAK490A	12 mL
• HRP Enzyme Catalogue Number MAK490B	120 μL
• LAA Enzyme Catalogue Number MAK490C	120 μL
• Dye Reagent Catalogue Number MAK490D	120 μL
• Standard (2 mM) Catalogue Number MAK490E	1 mL

Reagents and Equipment Required but Not Provided

- Pipetting devices and accessories (for example, multichannel pipettor, pipette tips, etc.)
- Multiwell plate reader.
- Clear flat-bottom 96-well plates for colorimetric assay or black flat-bottom 96-well plates for fluorometric assay. Cell culture or tissue culture treated plates are not recommended.
- 1.5 mL microcentrifuge tubes.
- Dounce tissue grinder set.
(Catalogue Number D9063 or equivalent)
- Refrigerated microcentrifuge capable of $\text{RCF} \geq 14,000 \times g$
- Phosphate Buffered Saline (PBS)
(Catalogue Number P3813 or equivalent)
- Potassium phosphate monobasic
(Catalogue Number P0662 or equivalent)

Precautions and Disclaimer

For R&D use only. Not for drug, household, or other uses. Please consult the Safety Data Sheet for information regarding hazards and safe handling practices.

Storage/Stability

The kit is shipped on wet ice. Store components at -20°C .

Preparation Instructions

Briefly centrifuge small vials prior to opening. Equilibrate all components to room temperature prior to use.

Procedure

All Samples and Standards should be run in duplicate.

Sample Preparation

Tissue or Solid Samples (for example, food)

1. Homogenize 20-100 mg of Sample in 200-1000 μL purified water.
2. Centrifuge at $10,000 \times g$ for 15 minutes at 4°C . Use the clear supernatant for the assay.

Cell Lysate

1. Collect cells by centrifugation at $2,000 \times g$ for 5 minutes at 4°C .
2. For adherent cells, do not harvest cells using proteolytic enzymes. Instead, use a rubber policeman or cell scraper.
3. Homogenize or sonicate cells in an appropriate volume of cold Assay Buffer containing 50 mM potassium phosphate (pH 7.5).
4. Centrifuge at $14,000 \times g$ for 10 minutes at 4°C . Use the clear supernatant for the assay.

Liquid Samples

Assay can be performed directly on liquid samples. It is recommended to dilute serum and cell culture media samples 4-fold in purified water prior to assay.

All Samples (Except Urine)

Transfer 20 μL of Sample to wells of a 96-well plate.

Urine Samples

1. Urine Samples must be diluted at least 5-fold with purified water.
2. Urine Samples require three separate reactions:
 - a. Sample

Sample Plus Internal Standard

- b. Sample Blank
3. For the Internal Standard, prepare a 1250 μM Standard by mixing 125 μL of 2 mM Standard with 75 μL of purified water.
 4. For the Sample Plus Internal Standard well, add 5 μL of the 1250 μM Internal Standard from Step 3 and 20 μL of Sample.
 5. For the Sample and Sample Blank wells, add 5 μL of purified water and 20 μL Sample.

Colorimetric Standard Curve Preparation

1. Prepare a 500 μM Standard by mixing 50 μL of the 2 mM Standard and 150 μL of purified water.
2. Prepare Standards in 1.5 mL microcentrifuge tubes according to Table 1.

Table 1.

Preparation of L-Amino Acid Colorimetric Standards

Well	500 μM Standard	Purified water	L-Amino Acid (μM)
1	100 μL	-	500
2	60 μL	40 μL	300
3	30 μL	70 μL	150
4	-	100 μL	0

3. Mix well and transfer 20 μL of each Standard into separate wells of a clear 96-well plate.

Fluorometric Standard Curve Preparation

1. Prepare Standards according to Colorimetric Standard Curve Preparation section.
2. Mix 10 μL of the Standards from Colorimetric Procedure with 90 μL of purified water according to Table 2.

Table 2.

Preparation of L-Amino Acid Fluorometric Standards

Well	Colorimetric Standard	Purified Water	L-Amino Acid (μM)
1	10 μL of 500 μM Std	90 μL	50
2	10 μL of 300 μM Std	90 μL	30
3	10 μL of 150 μM Std	90 μL	15
4	-	100 μL	0

3. Mix well and transfer 20 μL of each Standard into separate wells of a black 96-well plate.

Working Reagent

1. Mix enough reagent for the number of assays to be performed. For each well, prepare 88 µL of Working Reagent according to Table 3.
 - a. For urine Samples only, prepare 88 µL of Sample Blank Working Reagent for each Sample Blank well.

Table 3.

Preparation of Working Reagent

Reagent	Working Reagent	Sample Blank Working Reagent for Urine Samples
Assay Buffer	85 µL	86 µL
LAA Enzyme	1 µL	-
HRP Enzyme	1 µL	1 µL
Dye reagent	1 µL	1 µL

2. Transfer 80 µL of Working Reagent into each reaction well.
 - a. For urine samples, transfer 80 µL of Sample Blank Working Reagent to Sample Blank wells.
3. Tap plate to mix.

Measurement

1. Incubate at room temperature for 60 minutes. Protect plate from light for fluorometric assay.
2. Measure the optical density (OD) at 570 nm or fluorescence intensity (F) at $\lambda_{Ex} = 530 \text{ nm}/\lambda_{Em} = 585 \text{ nm}$.

Results

1. Calculate ΔOD or ΔF by subtracting the reading (OD or fluorescence intensity F) of Standard #4 (Blank) from the remaining Standard reading values.
2. Plot the ΔOD or ΔF against standard concentrations and determine the slope of the Standard curve.
3. Calculate the L-Amino Acid concentration of Samples using the below equation:
$$\text{L-Amino Acid } (\mu\text{M}) = \frac{R_{\text{Sample}} - R_{\text{Blank}}}{\text{Slope } (\mu\text{M}^{-1})} \times \text{DF}$$
4. For urine Samples, follow Steps 1-2 above and then calculate the L-Amino Acid concentration of Samples using the below equation:

Urine L-Amino Acid (μM) =

$$\frac{R_{\text{Sample}} - R_{\text{Sample Blank}}}{R_{\text{Standard}} - R_{\text{Sample}}} \times \frac{1250}{4} \times \text{DF}$$

where:

R_{Sample} = OD or fluorescence intensity (F) reading of Sample

R_{Blank} = OD or fluorescence intensity (F) reading of Blank

$R_{\text{Sample Blank}}$ = OD or fluorescence intensity (F) reading of Sample Blank (urine Samples only)

1250 = Concentration in μM of Internal Standard

4 = The volume of the Internal Standard is 4× lower than the Sample volume. In order to correct the calculation, the Internal Standard concentration is divided by 4.

DF = Sample dilution factor (DF = 1 for undiluted Samples)

Note: The Standard is 2 mM L-leucine, therefore the results are in terms of μM of L-leucine.

Figure 1.

Typical Colorimetric L-Amino Acids
Standard Curve





Figure 2.

Typical Fluorometric L-Amino Acids
Standard Curve



Notice

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

The information in this document is subject to change without notice and should not be construed as a commitment by the manufacturing or selling entity, or an affiliate. We assume no responsibility for any errors that may appear in this document.

Technical Assistance

Visit the tech service page at SigmaAldrich.com/techservice.

Terms and Conditions of Sale

Warranty, use restrictions, and other conditions of sale may be found at SigmaAldrich.com/terms.

Contact Information

For the location of the office nearest you, go to SigmaAldrich.com/offices

The life science business of Merck KGaA, Darmstadt, Germany
operates as MilliporeSigma in the U.S. and Canada.

MilliporeSigma, and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. All other trademarks are the property of their respective owners. Detailed information on trademarks is available via publicly accessible resources.

© 2022 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

mak490pis Rev 12/23

**MILLIPORE
SIGMA**