



Quality Control Ranges
Milliplex PLEXpediton Magnetic Bead Screening Panel
Catalog# HPLX1-115SP, HPLX1-115SP-PX, HPLX1-115SP-PX80
Control Catalog # HPLX1-6010-03 Lot # PLX13-101 and PLX13-201

<u>Bead</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>	<u>Bead</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
HBCA1	Control 1	138 - 286	pg/mL	HMCP2	Control 1	47 - 97	pg/mL
	Control 2	584 - 1212	pg/mL		Control 2	192 - 400	pg/mL
HGCP2	Control 1	18 - 38	pg/mL	HMIP1B	Control1	435 - 903	pg/ml
	Control 2	77 - 159	pg/mL		Control 2	1860 - 3862	pg/ml
HENA78	Control 1	45 - 94	pg/mL	HRANTES	Control1	27 - 56	pg/ml
	Control 2	190 - 395	pg/mL		Control 2	118 - 245	pg/ml
HEXTN2	Control 1	68 - 141	pg/mL	HITAC	Control1	18 - 37	pg/ml
	Control 2	294 - 611	pg/mL		Control 2	78 - 162	pg/ml
HFRAC TALKN	Control 1	971 - 2018	pg/mL	HMCP4	Control1	36 - 75	pg/ml
	Control 2	4285 - 8899	pg/mL		Control 2	156 - 323	pg/ml
HETXN3	Control 1	163 - 339	pg/mL	HMCP3	Control1	65 - 135	pg/ml
	Control 2	689 - 1431	pg/mL		Control 2	262 - 545	pg/ml
HGR0	Control 1	41 - 84	pg/mL	HMIP3A	Control1	12 - 25	pg/ml
	Control 2	187 - 387	pg/mL		Control 2	52 - 108	pg/ml
HI309	Control 1	13 - 28	pg/mL	HMIP1D	Control 1	466 - 968	pg/mL
	Control 2	57 - 119	pg/mL		Control 2	1895 - 3937	pg/mL
HETXN	Control 1	89 - 185	pg/mL	HTARC	Control1	36 - 75	pg/ml
	Control 2	374 - 776	pg/mL		Control 2	152 - 315	pg/ml
HMIG	Control 1	45 - 94	pg/mL	HMIP3B	Control 1	44 - 90	pg/mL
	Control 2	198 - 411	pg/mL		Control 2	178 - 369	pg/mL
HMIP1A	Control 1	173 - 359	pg/mL	HSDF1	Control1	196 - 407	pg/ml
	Control 2	754 - 1566	pg/mL		Control 2	821 - 1705	pg/ml

Note: The Quality Control Ranges are generated with overnight assay format using serum matrix provided in the kit.
HPLX1-6010-03: PLX13-101.201 (09 September 2025) EMD Millipore • (800) 645-5476 • (800) 645-5439 (FAX) • www.millipore.com