# CCL8 Hu-Cy5 SmartFlare™

**RNA Detection Probe** 

pack size: 50µL (250 rxns)

Cat. # SF-1498

Store at 2-8℃, after reconstitution store at 23-27℃ DO NOT FREEZE

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES. NOT FOR HUMAN OR ANIMAL CONSUMPTION



# **Product Data Sheet**

page 1 of 1

Validated Accession NM\_005623.2

Species Hu

Gene Aliases HC14, MCP2, MCP-2

### Confirmation of CCL8 SmartFlare Performance:

CCL8 SmartFlare has been tested in a buffer system to detect the release of the fluorophore in the presence of a complementary base pair sequence for each lot to confirm target specificity.

CCL8 SmartFlare has also been tested in a cell model system and demonstrated increased fluorescence in cells expressing the target compared to a scrambled negative control SmartFlare (Figure 1). For additional accession numbers predicted to react with this SmartFlare Probe. please visit http://www.millipore.com/catalogue/item/SF-1498




Figure 1: CCL8 Mean Fluorescence Intensity (green) measured by flow cytometry in living HUVEC cells demonstrated a significant increase over unflared cells (blue) as well as scramble control (red). Data shown in graph is representative.

### Storage and Handling:

Material has been 0.22µm filtered. Stable for 5 years at 2-8℃ degrees in lyophilized format ONLY. temperature is required for reconstituted product.

Warning-after reconstitution product is sensitive to cold and hot temperatures, a stable room temperature of 23-27℃ is required.

### **Handling Recommendations:**

Reconstitute with sterile nuclease free water in a drop wise fashion and tap tube repeatedly to fully dissolve lyophilized material. Vortex for 5-10 sec.

Upon reconstitution, store at room temperature for up to 1 year protected from light. Product must be handled with gloves as product can be absorbed through the skin.

## **Recommended Cell Testing Protocol:**

(example: 30.000 cells in a 200uL media volume within each well of a 96 well plate)

- Reconstitute reagent in 50µL of sterile nuclease free
- Create a working solution based on your experiment by diluting 1:20 in sterile PBS.
- Add 4µL directly to cells (at approx 80% confluency).
- Allow to incubate overnight for 16 hrs.
- Detect using fluorescence detection platform of choice.

