



SIGMA QUALITY CONTROL TEST PROCEDURE

Product Information

Enzymatic Assay of ACYLASE I¹ (EC 3.5.1.14) from Porcine

PRINCIPLE:

N-Acetyl-L-Methionine + H₂O $\xrightarrow{\text{Acylase I}}$ L-Methionine + Acetic Acid

CONDITIONS: T = 25°C, pH = 7.0, A_{238nm}, Light path = 1 cm

METHOD: Continuous Spectrophotometric Rate Determination

REAGENTS:

- A. 100 mM Potassium Phosphate Buffer, pH 7.0 at 25°C
(Prepare 100 ml in deionized water using Potassium Phosphate, Monobasic, Anhydrous, Sigma Prod. No. P-5379. Adjust to pH 7.0 at 25°C with 1 M KOH.)
- B. 15 mM N-Acetyl-L-Methionine Solution (NAMet)
(Prepare 20 ml in Reagent A using N-Acetyl-L-Methionine, Sigma Prod. No. A-3258.)
- C. Acylase I Enzyme Solution
(Immediately before use, prepare a solution containing 1500 - 2500 units/ml of Acylase I in Reagent A. Let stand for 5-10 minutes at room temperature prior to assaying)

PROCEDURE:

Pipette (in milliliters) the following reagents into suitable quartz cuvettes:

	<u>Test</u>	<u>Blank</u>
Reagent B (NAMet)	2.90	2.90

Enzymatic Assay of ACYLASE I¹
(EC 3.5.1.14)
from Porcine

PROCEDURE: (continued)

Equilibrate to 25°C. Monitor the $A_{238\text{nm}}$ until constant, using a suitably thermostatted spectrophotometer. Then add:

	<u>Test</u>	<u>Blank</u>
Reagent A (Buffer)	-----	0.10
Reagent C (Enzyme Solution)	0.10	-----

Immediately mix by inversion and record the decrease in $A_{238\text{nm}}$ for approximately 5 minutes. Obtain the $\Delta A_{238\text{nm}}/\text{minute}^2$ using the maximum linear rate for both the Test and Blank.

CALCULATIONS:

$$\text{Units/ml enzyme} = \frac{(\Delta A_{238\text{nm}}/\text{min Test} - \Delta A_{238\text{nm}}/\text{min Blank})(60)(3)(\text{df})}{(0.019)(0.1)}$$

60 = Conversion factor from minutes to one hour as per the Unit Definition

3 = Total Volume (in milliliters) of assay

df = Dilution factor

0.019 = Millimolar extinction coefficient³ of N-Acetyl-L-Methionine at 238 nm

0.1 = Volume (in milliliter) of enzyme used

$$\text{Units/mg solid} = \frac{\text{units/ml enzyme}}{\text{mg solid/ml enzyme}}$$

$$\text{Units/mg protein} = \frac{\text{units/ml enzyme}}{\text{mg protein/ml enzyme}}$$

UNIT DEFINITION:

One unit will hydrolyze 1.0 μmole of N-acetyl-L-methionine per hour at pH 7.0 at 25°C.

FINAL ASSAY CONCENTRATION:

In a 3.00 ml reaction mix, the final concentrations are 100 mM potassium phosphate, 14 mM N-acetyl-L-methionine and 150 - 250 units acylase I.

REFERENCE:

Mitz, M.A. and Schlueter, R.J. (1958) *Biochimica Et Biophysica Acta* **27**, 168-172

Enzymatic Assay of ACYLASE I¹
(EC 3.5.1.14)
from Porcine

NOTES:

1. This assay is not to be used to assay Acylase I, from *Aspergillus melleus*, Sigma Prod. No. A-2156.
2. The $\Delta A_{238\text{nm}}$ /minute should not be over 0.05.
3. This value was determined by Sigma.
4. This assay is based on the cited reference.
5. Where Sigma Product or Stock numbers are specified, equivalent reagents may be substituted.

Sigma warrants that the above procedure information is currently utilized at Sigma and that all Sigma-Aldrich, Inc. products conform to the information in this and other Sigma-Aldrich, Inc. publications. Purchaser must determine the suitability of the information and product(s) for their particular use. Additional terms and conditions may apply. Please see reverse side of the invoice or packaging slip.