# Supelco.

## 1.01749.0001 1.01749.0007

# **Spectroquant® Volatile Organic Acids Cell Test**

## 1. Method

In an acidic medium lower fatty acids ("volatile organic acids") react with a diole to form fatty acid esters, which are subsequently converted into hydroxamic acids with hydroxylamine. These in turn react with iron(III) ions to form red complexes that are determined photometrically.

## 2. Measuring range and number of determinations

Measuring range	Number of determinations
50 - 3000 mg/l <sup>1)</sup> 71 - 4401 mg/l <sup>2)</sup>	25

<sup>1)</sup> calculated as acetic acid <sup>2)</sup> calculated as butyric acid

For programming data for selected photometers / spectrophotometers see www.service-test-kits.com.

## 3. Applications

Sample material: Digested sludge Activated sludge Process water

## 4. Influence of foreign substances

This was checked individually in solutions containing 1500 and 0 mg/l acetic acid. The determination is not yet interfered with up to the concentrations of foreign substances given in the table. Cumulative effects were not checked; such effects can, however, not be excluded.

Concentrations of foreign substances in mg/l or %					
$\begin{array}{c} AI^{3+}\\ Ca^{2+}\\ Cd^{2+}\\ Cr^{3+}\\ Cu^{2+}\\ Fe^{3+}\\ Hg^{2+}\\ Mg^{2+}\\ \end{array}$	1000 1000 50 50 1000 50 1000	$\begin{array}{c} Mn^{2+} \\ NH_4^+ \\ Ni^{2+} \\ Pb^{2+} \\ PO_4^{3^-} \\ Zn^{2+} \end{array}$	1000 50 50 1000	Acetaldehyde Acetone Ethanol Formaldehyde Surfactants <sup>1)</sup> NaCl NaNO <sub>3</sub> Na>SO <sub>4</sub>	50 1000 5 % 50 1000 20 % 20 % 10 %

<sup>1)</sup> tested with nonionic, cationic, and anionic surfactants

## 5. Reagents and auxiliaries

#### Please note the warnings on the packaging materials!

The test reagents are stable up to the date stated on the pack when stored closed at +15 to +25 °C.

Package contents:

- 1 bottle of reagent OA-1K
- 1 bottle of reagent OA-2K 1 bottle of reagent OA-3K
- 1 bottle of reagent OA-4K
- 25 reaction cells
- 1 sheet of round stickers for numbering the cells

#### Other reagents and accessories:

MQuant<sup>®</sup> Universal indicator strips pH 0 - 14, Cat. No. 109535

Pipettes for pipetting volumes of 0.50 and 5.0 ml Thermoreactor

#### 6. Preparation

- Analyze immediately after sampling. Otherwise store at 4 °C for max. 24 hours.
- The pH must be within the range 2 12.
- · Filter turbid samples.

## 7. Procedure

	0.50	<b>B</b> 1 11 11 11		
Reagent OA-1K	0.50 ml	Pipette into a reaction cell.		
Pretreated sample	0.50 ml	Add with pipette, close the cell tightly, and mix.		
Heat the cell at 100 °C in the preheated thermoreactor for 15 min, then cool to room temperature under running water.				
Reagent OA-2K	1.0 ml	Add with pipette.		
Reagent OA-3K and mix.	1.0 ml	Add with pipette, close the cell tightly,		
Reagent OA-4K	1.0 ml	Add with pipette, close the cell tightly, and mix. A transient turbidity or precipitate may form.		
Leave to stand for 1 min (reaction time), then measure the sample in the				

#### Notes on the measurement:

photometer.

- For photometric measurement the cells must be clean. Wipe, if necessary, with a clean dry cloth.
- Measurement of turbid solutions yields false-high readings.
- The color of the measurement solution remains stable for 30 min after the end of the reaction time stated above.

## 8. Analytical quality assurance

recommended before each measurement series To check the photometric measurement system (test reagents, measure-ment device, handling) and the mode of working, a freshly prepared standard solution containing 2.05 g/l of sodium acetate anhydrous (corresponding to 1500 mg/l of acetic acid) (application see the website) can be used.

#### Sample-dependent interferences (matrix effects) can be determined by means of standard addition.

Additional notes see under www.qa-test-kits.com. For quality and batch certificates for Spectroquant<sup>®</sup> test kits see the website, where you will find all data in production control, that are determined in accordance with ISO 8466-1 and DIN 38402 A51.

### 9. Notes

- Reclose the reagent bottles immediately after use.
- Information on disposal can be obtained at www.disposal-test-kits.com.

