

QUALITY CONTROL RANGES



MILLIPLEX[®] MAP Human Neurological Disorders Magnetic Bead Panel 3
Kit Catalog # HND3MAG-39K
Control Catalog # HND3-6039-3
Lot# HND3-106 and HND3-206

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
SOD1	Control I	0.36 - 0.75	ng/mL
	Control II	3.2 - 6.6	ng/mL
Angiotensinogen	Control I	12 - 25	ng/mL
	Control II	104 - 216	ng/mL
Fetuin A	Control I	45 - 93	ng/mL
	Control II	372 - 772	ng/mL
SOD2	Control I	0.29 - 0.59	ng/mL
	Control II	2.8 - 5.8	ng/mL
Contactin-1	Control I	0.72 - 1.5	ng/mL
	Control II	6.1 - 13	ng/mL
Kallikrein-6	Control I	0.35 - 0.73	ng/mL
	Control II	3.1 - 6.5	ng/mL
Osteopontin	Control I	1.3 - 2.7	ng/mL
	Control II	11 - 23	ng/mL

**Note: The Quality Control Ranges are generated with overnight assay format.
Quality Control values in culture media are not tested.**

HND3-6039-3: HND3-106.206 (30 September 2022)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA