

## User Manual

# CaspaTag™ Caspase-8 In Situ Assay Kit

**APT428**
**Pack Size 25 Assays**
**Store at 2-8 °C**
**FOR RESEARCH USE ONLY**
**Not for use in diagnostic procedures. Not for Human or Animal Consumption.**

## Background

Apoptosis is an evolutionarily conserved form of cell suicide, which follows a specialized cellular process. The central component of this process is a cascade of proteolytic enzymes called caspases. These enzymes participate in a series of reactions that are triggered in response to pro-apoptotic signals and result in the cleavage of protein substrates, causing the disassembly of the cell.<sup>1</sup>

Caspases have been identified in organisms ranging from *C. elegans* to humans. The mammalian caspases play distinct roles in apoptosis and inflammation. In apoptosis, caspases are responsible for proteolytic cleavages that lead to cell disassembly (effector caspases) and are involved in upstream regulatory events (initiator caspases). An active caspase consists of two large and two small subunits that form two heterodimers, which associate in a tetramer.<sup>2-4</sup> In common with other proteases, caspases are synthesized as precursors that undergo proteolytic maturation, either autocatalytically or in a cascade by enzymes with similar specificity.<sup>5</sup>

Caspase enzymes specifically recognize a 4 or 5 amino acid sequence on the target substrate, which necessarily includes an aspartic acid residue. This residue is the target for cleavage, which occurs at the carbonyl end of the aspartic acid residue.<sup>6</sup> Caspases can be detected via immunoprecipitation, immunoblotting techniques using caspase specific antibodies, or by employing fluorochrome substrates, which become fluorescent upon cleavage by the caspase.

## Test Principle

CHEMICON®'s CaspaTag™ Caspase-8 In Situ Assay Kits use a novel approach to detect active caspases. The methodology is based on Fluorochrome Inhibitors of Caspases (FLICA). The inhibitors are cell permeable and non-cytotoxic. Once inside the cell, the inhibitor binds covalently to the active caspase.<sup>7</sup> This kit uses a carboxyfluorescein-labeled fluoromethyl ketone peptide inhibitor of caspase-8 (FAM-LETD-FMK), which produces a green fluorescence. When added to a population of cells, the FAM-LETD-FMK probe enters each cell and covalently binds to a reactive cysteine residue that resides on the large subunit of the active caspase heterodimer, thereby inhibiting further enzymatic activity. The bound labeled reagent is retained within the cell, while any unbound reagent will diffuse out of the cell and is washed away. The green, fluorescent signal is a direct measure of the amount of active caspase-8 present in the cell at the time the reagent was added. Cells that contain the bound labeled reagent can be analyzed by 96-well plate-based fluorometry, fluorescence microscopy, or flow cytometry.

## Application

The CHEMICON® CaspaTag™ Caspase-8 In Situ Assay Kit, Fluorescein is a fluorescent-based assay for detection of active caspase-8 in cells undergoing apoptosis. The kit is for research use only. Not for use in diagnostic or therapeutic procedures.

## Kit Components

1. FLICA Reagent (FAM-LETD-FMK): One lyophilized vial
2. 10X Wash Buffer: 15 mL
3. Fixative: 6 mL
4. Propidium Iodide: 1 mL at 250 µg/mL, ready-to-use
5. Hoechst Stain: 1 mL at 200 µg/mL, ready-to-use

## Materials Required

- Cultured cells with media
- Reagents to induce apoptosis
- 15 mL polystyrene centrifuge tubes
- Amber vials or tubes for storage of 150X concentrate at –20 °C
- 600 mL graduated cylinder
- Microscope slides
- Hemocytometer
- Centrifuge (400 x g)
- 37 °C incubator
- Vortexer
- Adjustable volume pipettor with disposable tips
- Deionized water
- PBS, pH 7.4
- DMSO

## Warnings and Precautions

- Use gloves, protective clothing and eyewear when handling the FLICA reagent, Propidium iodide, Hoechst stain and fixative.
- Dispose of all liquid components down the sink and flush with copious amounts of water. Solid components may be tossed in standard trash bins.

## Storage and Stability

- Store unopened kit materials at 2-8 °C up to their expiration date.
- Reconstituted FLICA Reagent (150X) should be frozen at –20 °C for up to 6 months and may be thawed twice during this time. Aliquot into separate amber tubes if desired. Protect from light at all times.
- Store diluted (1X) wash buffer up to –20 °C for 2 weeks.

## Preparation of Reagents

1. Wash buffer: If necessary, gently warm the 10X concentrate supplied in the kit to completely dissolve any salt crystals that have come out of solution. Prepare a 1X solution by diluting 1:10 in DI H<sub>2</sub>O.
2. FLICA Reagent (immediately before performing assay): Reconstitute one vial of lyophilized reagent with 50 µL of DMSO and mix by swirling or tilting the vial until completely dissolved (a few minutes at RT). This 150X stock solution should be used or frozen immediately after preparation. If using immediately, prepare a 1:5 dilution in PBS (pH 7.4), mixing well by vortex, to create a working 30X dilution. ALWAYS protect from light.

## Preparation of Samples

1. Cells should be cultured to a density optimal for apoptosis induction, not to exceed  $10^6$  cells/mL.
2. Induce apoptosis according to your protocol (camptothecin or staurosporine can be used for treatment for 3-4 hours).
3. Concurrently incubate a non-induced negative control cell population at the same density.
4. It is recommended that induced and non-induced samples be assayed for each labeling condition. For example, if using both the FLICA reagent and the Hoechst stain, make 8 populations.
  - Unlabeled, induced and non-induced
  - FLICA-labeled, induced and non-induced
  - FLICA and Hoechst-labeled, induced and non-induced
  - Hoechst-labeled, induced and non-induced

## Assay Instructions

### For Fluorescent Plate Reader

1. Concentrate both induced and non-induced cells to  $\sim 1 \times 10^7$  cells/mL.
2. Transfer 290-300  $\mu$ L of each cell suspension to sterile tubes.
3. Add 10  $\mu$ L of freshly prepared 30X FLICA reagent and mix cells by slightly flicking the tubes.  
**Note:** End user should adjust the amount of FLICA reagent used based on cell type and research conditions.
4. Incubate tubes for 1 hour at 37 °C under 5% CO<sub>2</sub>, protecting tubes from light. Swirl tubes once or twice during this time to gently resuspend settled cells.
5. Add 2 mL of 1X wash buffer to each tube and mix.
6. Centrifuge cells at  $< 400 \times g$  for 5 minutes at room temperature.
7. Carefully remove and discard supernatant, and gently vortex cell pellet to disrupt any cell-to-cell clumping.
8. Repeat wash (steps 5-7), this time using 1 mL of 1X wash buffer.
9. Resuspend cell pellet in 1 mL of 1X wash buffer.
10. Remove 50  $\mu$ L from each tube and dilute 1:10 with 450  $\mu$ L of PBS. Count cells with a hemocytometer or other method and compare density of induced and non-induced populations.
11. Centrifuge remaining cells from step 9 at  $< 400 \times g$  for 5 minutes at room temperature; remove and discard supernatant.
12. Resuspend non-induced cells in 400  $\mu$ L PBS.
13. Induced cell populations may have decreased due to loss of cells during apoptosis.
  - If populations of induced cells (based on count from step 10) were close to that of non-induced, resuspend *induced cells* in 400  $\mu$ L of PBS.
  - If there was a dramatic loss in induced cells counted in step 10, resuspend *induced cells* in PBS sufficient to adjust the concentration to that of the non-induced population ( $< 400 \mu$ L).
14. Place 100  $\mu$ L of each cell suspension into each of two wells of a black microtiter plate. Avoid creation of bubbles.
15. Set the fluorescence plate reader to perform an endpoint reading.
16. Read absorbance of each microwell using an excitation wavelength of 490 nm and an emission wavelength of 520 nm.

## For Fluorescence Microscopy- Adherent Cells

1. Trypsinize cells.
2. Count cells, and seed  $10^4$ - $10^5$  cells onto a sterile glass coverslip in a 35 mm petri dish or onto a chamber slide.
3. Grow cells in their respective cell culture media formulation for 24 hours at 37 °C.
4. Induce apoptosis using desired method.
5. Add freshly prepared 30X FLICA reagent solution at a 1:30 dilution in culture medium (10 µL in 290 µL of medium) and mix well.  
**Note:** End user should adjust the amount of FLICA reagent used based on cell type and research conditions.
6. Incubate for 1 hour at 37 °C under 5% CO<sub>2</sub> and remove medium.
7. Hoechst stain can be used to label the nuclei of dying cells. If this is desired, dilute 1.5 µL Hoechst stain in 300 µL of media (0.5% v/v). Add this media to cells and incubate for 5 minutes at 37 °C under 5% CO<sub>2</sub>.
8. Wash cells twice with 2 mL 1X wash buffer.
9. To analyze cells immediately, skip to step 10. However, if desired cells can be fixed at this point for analysis up to 24 hours later.
  - Dilute Fixative 1:10 with wash buffer.
  - Mount a cover slip with cells facing down onto a slide containing a drop of diluted fixative. If using chamber slides remove plastic frame, add a drop of diluted fixative onto slide and cover with a cover slip.
  - Keep fixed cells at 2-8 °C protected from light up to 24 hours. Proceed to step 11.
10. Mount a cover slip with cells facing down onto a slide containing a drop of 1X wash buffer. If using chamber slides remove plastic frame, add a drop of 1X wash buffer onto slide and cover with a cover slip.
11. Observe cells under a fluorescence microscope using a bandpass filter (excitation 490 nm, emission 520 nm) to view the green fluorescence of caspase-positive cells. Hoechst stain (if used) can be seen using a UV-filter with excitation at 365 nm and emission at 480 nm.

## For Fluorescent Microscopy- Suspension Cells

1. Transfer 290-300 µL of each cell suspension to sterile tubes.
2. Add 10 µL of freshly prepared 30X FLICA reagent and mix cells by slightly flicking the tubes.  
**Note:** End user should adjust the amount of FLICA reagent used based on cell type and research conditions.
3. Incubate tubes for 1 hour at 37 °C under 5% CO<sub>2</sub>, protecting tubes from light. Swirl tubes once or twice during this time to gently resuspend settled cells.
4. Hoechst stain can be used to label the nuclei of dying cells. If this is desired, add 1.5 µL Hoechst stain (0.5% v/v). Incubate for 5 minutes at 37 °C under 5% CO<sub>2</sub>.
5. Add 2 mL of 1X wash buffer to each tube and gently mix.
6. Centrifuge the cells at  $< 400 \times g$  for 5 minutes at room temperature.
7. Carefully remove and discard supernatant, and gently vortex cell pellet to disrupt any cell-to-cell clumping.
8. Repeat wash (steps 5-7), this time using 1 mL of 1X wash buffer.
9. Resuspend cell pellet in 300 µL of 1X wash buffer and place cells on ice.
10. Propidium Iodide can be used to exclude dead cells from the analysis. If desired, add 1.5 µL of PI solution (0.5% v/v). Cells may then be viewed using a long pass filter with excitation at 490 nm, emission  $> 520$ nm. PI has a maximum emission at 637 nm.

11. To analyze cells immediately, skip to step 12. However, if desired cells can be fixed at this point for analysis up to 24 hours later.
  - Add 30  $\mu$ L of fixative to each tube.
  - Incubate cells for 15 minutes at room temperature in the dark.
  - Dry cells onto a microscope slide.
  - Briefly wash cells with PBS.
  - Cover cells with mounting media and cover slip.
  - Keep fixed cells at 2-8  $^{\circ}$ C protected from light up to 24 hours. Proceed to step 13.
12. Place one drop of cell suspension onto a microscope slide and cover with a cover slip.
13. Observe cells under a fluorescence microscope using a bandpass filter (excitation 490 nm, emission 520 nm) to view the green fluorescence of caspase-positive cells. Hoechst stain (if used) can be seen using a UV-filter with excitation at 365 nm and emission at 480 nm.

### For Flow Cytometry

Flow Cytometry analysis can be done with single color (FLICA alone) or dual color staining (FLICA and Propidium Iodide). It is recommended that induced and non-induced samples be run for each labeling condition (unlabeled, FLICA-labeled, PI-labeled, and FLICA/PI-labeled).

1. Transfer 290-300  $\mu$ L of each cell suspension ( $\sim$  10<sup>6</sup> cells) to sterile tubes.
2. Add 10  $\mu$ L of freshly prepared 30X FLICA reagent and mix cells by slightly flicking the tubes.  
**Note:** End user should adjust the amount of FLICA reagent used based on cell type and research conditions.
3. Incubate tubes for 1 hour at 37  $^{\circ}$ C under 5% CO<sub>2</sub>, protecting tubes from light. Swirl tubes once or twice during this time to gently resuspend settled cells.
4. Add 2 mL of 1X wash buffer to each tube and gently mix.
5. Centrifuge the cells at  $< 400 \times g$  for 5 minutes at room temperature.
6. Carefully remove and discard supernatant, and gently vortex cell pellet to disrupt any cell-to-cell clumping.
7. Repeat wash (steps 4-6), this time using 1 mL of 1X wash buffer.
8. Resuspend cell pellet in 400  $\mu$ L of 1X wash buffer.
9. If using bi-color analysis, add 2  $\mu$ L of PI solution to one cell suspension. Set aside a second suspension without PI. Skip to step 12.
10. For immediate single-color analysis, put samples on ice and skip to step 11. For single-color analysis up to 24 hours later, add 40  $\mu$ L fixative and mix. Keep fixed cells at 2-8  $^{\circ}$ C protected from light.  
**NOTE:** Cells to be analyzed with PI cannot be fixed.
11. For single-color analysis, use a 15 mW argon ion laser at 488 nm. Measure fluorescein on the FL1 channel. Generate a log FL1 (X-axis) versus number of cells (Y-axis) histogram. On the histogram, there will appear two cell populations represented by two peaks. The majority of the caspase negative (-) cells will normally occur within the first log decade of the FL1 (X) axis (first peak), whereas the caspase-positive (+) cell population will appear as a separate peak or as a shoulder of the first peak showing increased fluorescence intensity. Position the vertical cursor in the gap between the two peaks. Events falling to the right of the vertical cursor should be counted as caspase positive (+).
12. For bicolor analysis, measure fluorescein on the FL1 channel and red fluorescence (PI) on the FL2 channel. Generate a log FL1 (X-axis) versus log FL2 (Y-axis) dot plot. Put in quadrant cursors. The 4 quadrant areas contain the following cell populations: (i) quadrant 1, PI positive fluorescein negative cells; (ii) quadrant 2, fluorescein positive PI positive cells; (iii) quadrant 3, fluorescein negative PI negative cells; (iv) quadrant 4, fluorescein positive PI negative cells. The cell population in quadrant 4 consists of living caspase positive (+) cells.

## References

1. Slee E.A. et al. (1999). *Cell Death and Differ.* **6**: 1067-1074.
2. Walker N.P. et al. (1994). *Cell* **78**: 343-352.
3. Wilson K.P. et al. (1994). *Nature* **370**: 270-275.
4. Rotonda J. et al. (1996). *Nature Struct. Biol.* **3(7)**: 619-625.
5. Kumar S. (1999). *Cell Death and Differ.* **6**: 1060-1066.
6. Thornberry N.A. et al. (1997). *J. Biol. Chem.* **272(29)**: 17907-17911.
7. Ekert P.G. et al. (1999). *Cell Death and Differ.* **6**: 1081-1086.

## Notice

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

The information in this document is subject to change without notice and should not be construed as a commitment by the manufacturing or selling entity, or an affiliate. We assume no responsibility for any errors that may appear in this document.

## Technical Assistance

Visit the tech service page at [SigmaAldrich.com/techservice](https://SigmaAldrich.com/techservice).

## Terms and Conditions of Sale

Warranty, use restrictions, and other conditions of sale may be found at [SigmaAldrich.com/terms](https://SigmaAldrich.com/terms).

## Contact Information

For the location of the office nearest you, go to [SigmaAldrich.com/offices](https://SigmaAldrich.com/offices).

The life science business of Merck KGaA, Darmstadt, Germany  
operates as MilliporeSigma in the U.S. and Canada.

MilliporeSigma, Chemicon Millipore and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates. All other trademarks are the property of their respective owners. Detailed information on trademarks is available via publicly accessible resources.  
© 2023 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

Document Template 00007539FM, Ver 13.0

APT428 Ver 4.0, Rev 26JUL2023, NM

