

## SIGMA QUALITY CONTROL TEST PROCEDURE

**Product Information****Enzymatic Assay of HEXOKINASE<sup>1</sup>**  
**(EC 2.7.1.1)****PRINCIPLE:**

D-Glucose + ATP  $\xrightarrow{\text{Hexokinase}}$  D-Glucose 6-Phosphate + ADP

D-Glucose 6-Phosphate +  $\beta$ -NADP  $\xrightarrow{\text{G-6-PDH}}$  6-PG +  $\beta$ -NADPH

Abbreviations used:

ATP = Adenosine 5'-Triphosphate

ADP = Adenosine 5'-Diphosphate

G-6-PDH = Glucose-6-Phosphate Dehydrogenase

$\beta$ -NADP =  $\beta$ -Nicotinamide Adenine Dinucleotide Phosphate, Oxidized Form

$\beta$ -NADPH =  $\beta$ -Nicotinamide Adenine Dinucleotide Phosphate, Reduced Form

6-PG = 6-Phospho-D-Gluconate

**CONDITIONS:** T = 25°C, pH = 7.6, A<sub>340nm</sub>, Light path = 1 cm

**METHOD:** Continuous Spectrophotometric Rate Determination

**REAGENTS:**

- A. 50 mM Triethanolamine Buffer, pH 7.6 at 25°C  
(Prepare 100 ml in deionized water using Triethanolamine Hydrochloride, Sigma Prod. No. T-1502. Adjust to pH 7.6 at 25°C with 1 M NaOH.)
- B. 555 mM D-Glucose Solution  
(Prepare 10 ml in Reagent A using D-(+)-Glucose, Anhydrous, Sigma Prod. No. G-8270.)
- C. 19 mM Adenosine 5'-Triphosphate Solution (ATP)  
(Prepare 10 ml in deionized water using Adenosine 5'-Triphosphate, Disodium Salt, Sigma Prod. No. A-2383. **PREPARE FRESH.**)
- D. 100 mM Magnesium Chloride Solution (MgCl<sub>2</sub>)  
(Prepare 5 ml in deionized water using Magnesium Chloride, Hexahydrate, Sigma Prod. No. M-0250.)

# **Enzymatic Assay of HEXOKINASE<sup>1</sup>** **(EC 2.7.1.1)**

## **REAGENTS:** (continued)

- E. 14 mM  $\beta$ -Nicotinamide Adenine Dinucleotide Phosphate, Oxidized Form, Solution ( $\beta$ -NADP)  
 (Dissolve the contents of two 10 mg vials of  $\beta$ -Nicotinamide Adenine Dinucleotide Phosphate, Sodium Salt, Sigma Stock No. 240-310, in the appropriate volume of deionized water **or** prepare 10 ml in deionized water using  $\beta$ -Nicotinamide Adenine Dinucleotide Phosphate, Sodium Salt, Sigma Prod. No. N-0505. **PREPARE FRESH.**)
- F. Glucose-6-Phosphate Dehydrogenase Enzyme Solution (G-6-PDH)<sup>2</sup>  
 (Immediately before use, prepare a solution containing 125 units/ml of Glucose-6-Phosphate Dehydrogenase, Sigma Prod. No. G-4134, in cold Reagent A.)<sup>3</sup>
- G. Hexokinase Enzyme Solution  
 (Immediately before use, prepare a solution containing 0.5 - 1.0 unit/ml of Hexokinase in cold deionized water.)

## **PROCEDURE:**

Pipette (in milliliters) the following reagents into suitable cuvettes:

|                                | <u>Test</u> | <u>Blank</u> |
|--------------------------------|-------------|--------------|
| Reagent A (Buffer)             | 1.00        | 1.00         |
| Reagent B (D-Glucose)          | 1.00        | 1.00         |
| Reagent C (ATP)                | 0.10        | 0.10         |
| Reagent D (MgCl <sub>2</sub> ) | 0.20        | 0.20         |
| Reagent E ( $\beta$ -NADP)     | 0.20        | 0.20         |
| Reagent F (G-6-PDH)            | 0.02        | 0.02         |

Mix by inversion and equilibrate to 25°C. Monitor the  $A_{340\text{nm}}$  until constant, using a suitably thermostatted spectrophotometer. Then add:

|                             |       |       |
|-----------------------------|-------|-------|
| Deionized Water             | ----- | 0.05  |
| Reagent G (Enzyme Solution) | 0.05  | ----- |

Immediately mix by inversion and record the increase in  $A_{340\text{nm}}$  for approximately 5 minutes. Obtain the  $\Delta A_{340\text{nm}}/\text{minute}$  using the maximum linear rate for both the Test and Blank.

## Enzymatic Assay of HEXOKINASE<sup>1</sup> (EC 2.7.1.1)

### CALCULATIONS:

$$\text{Units/ml enzyme} = \frac{(\Delta A_{340\text{nm}}/\text{min Test} - \Delta A_{340\text{nm}}/\text{min Blank})(2.57)(\text{df})}{(6.22)(0.05)}$$

2.57 = Total volume (in milliliters) of assay

df = Dilution factor

6.22 = Millimolar extinction coefficient of  $\beta$ -NADPH at 340 nm

0.05 = Volume (in milliliter) of enzyme used

$$\text{Units/mg solid} = \frac{\text{units/ml enzyme}}{\text{mg solid/ml enzyme}}$$

$$\text{Units/mg protein} = \frac{\text{units/ml enzyme}}{\text{mg protein/ml enzyme}}$$

### UNIT DEFINITION:

One unit will phosphorylate 1.0  $\mu$ mole of D-glucose per minute at pH 7.6 at 25°C.

### FINAL ASSAY CONCENTRATION:

In a 2.57 ml reaction mix, the final concentrations are 39 mM triethanolamine, 216 mM D-glucose, 0.74 mM adenosine 5'-triphosphate, 7.8 mM magnesium chloride, 1.1 mM  $\beta$ -nicotinamide adenine dinucleotide phosphate, 2.5 units glucose-6-phosphate dehydrogenase, and 0.025 - 0.05 unit of hexokinase.

### REFERENCES:

Bergmeyer, H.U., Grassl, M., and Walter, H.E. (1983) in *Methods of Enzymatic Analysis* (Bergmeyer, H.U. ed) 3rd ed., Volume II, 222-223, Verlag Chemie, Deerfield Beach, FL

### NOTES:

1. This procedure is not to be used to assay the activity of Hexokinase, Sigma Prod. No. H-3779, Hexokinase, Insoluble enzyme attached to beaded agarose, Sigma Prod. No. H-2005, and Hexokinase, Insoluble enzyme attached to polyacrylamide, Sigma Prod. No. H-8254.

**Enzymatic Assay of HEXOKINASE<sup>1</sup>**  
**(EC 2.7.1.1)**

**NOTES**

2. Glucose-6-Phosphate Dehydrogenase unit definition: One unit will oxidize 1.0  $\mu$ mole of D-glucose 6-phosphate to 6-phospho-D-gluconate per minute in the presence of  $\beta$ -NADP at pH 7.4 at 25°C.
3. Other types of glucose-6-phosphate dehydrogenase may contain varying amounts of hexokinase as an impurity.
4. Where Sigma Product or Stock numbers are specified, equivalent reagents may be substituted.

**Sigma warrants that the above procedure information is currently utilized at Sigma and that Sigma products conform to the information in Sigma publications. Purchaser must determine the suitability of the information and products for its particular use. Upon purchase of Sigma products, see reverse side of invoice or packing slip for additional terms and conditions of sale.**