

## Quality Control Ranges



### MILLIPLEX<sup>®</sup> MAP Sepsis Magnetic Bead Panel 2 Kit Catalog # HSP2MAG-63K

#### Control Catalog # HSP2-6063-2, Lot# HSP2-108 and HSP2-208

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
Granzyme B	Control I	18 - 37	pg/mL
	Control II	94 - 195	pg/mL
HSP70	Control I	25739 - 53458	pg/mL
	Control II	132101 - 274364	pg/mL
IL-1 $\alpha$	Control I	87 - 180	pg/mL
	Control II	465 - 965	pg/mL
IL-8	Control I	42 - 87	pg/mL
	Control II	223 - 463	pg/mL
MIP-1 $\alpha$	Control I	81 - 168	pg/mL
	Control II	329 - 683	pg/mL
MIP-1 $\beta$	Control I	85 - 178	pg/mL
	Control II	429 - 890	pg/mL
MMP-8	Control I	994 - 2065	pg/mL
	Control II	5743 - 11928	pg/mL

**Note: The Quality Control Ranges are generated with overnight assay format using serum matrix provided in the kit. Quality Control values in culture media are not tested.**

HSP2-6063-2: HSP2-108.208 (20 June 2022)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • [emdmillipore.com](http://emdmillipore.com)

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA