



QUALITY CONTROL RANGES
MILLIPLEX[®] MAP Human MMP Magnetic Bead Panel 1
Kit Catalog # HMMP1MAG-55K

Control Catalog # HMMP1MG-6055-1
Lot# HMMPM-111 and HMMPM-211

<u>Hormone</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
HMMP-3	Control I	852 - 1770	pg/mL
	Control II	8056 - 16733	pg/mL
HMMP-12	Control I	701 - 1455	pg/mL
	Control II	5175 - 10748	pg/mL
HMMP-13	Control I	344 - 715	pg/mL
	Control II	3066 - 6368	pg/mL

Note: The Quality Control Ranges are generated with same day assay format using serum matrix provided in the kit. Quality Control values in culture media are not tested.

HMMP1MG-6055-1: HMMPM-111.211 (20 August 2024)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA