

## EC Declaration of Conformity

Application of Council Directive(s) 2004/108/EC  
2006/95/EC using EN61010-1

Manufacture's Name: Glas-Col LLC  
Manufacture's Address: 711 Hulman Street  
Terre Haute, IN 47802 U.S.A.

*Declares that the product*

Product Name: Temperature Control  
Model Numbers: 104A PL624CE, KCE, TCE

*Conforms to the following Product Specifications:*


### **2004/108/EC Electromagnetic Compatibility Directive**

<b>EN 61326-1</b>	<b>2006</b>		<b>Electrical equipment for measurement, control and laboratory use – EMC requirements (Industrial Immunity, Class B Emissions).</b>
EN 61000-4-2	1996	A1, A2, 2001	Electrostatic Discharge Immunity
EN 61000-4-3	2006		Radiated Field Immunity
EN 61000-4-4	2004		Electrical Fast-Transient / Burst Immunity
EN 61000-4-5	2006		Surge Immunity
EN 61000-4-6	1996	A1, 2, 3, 2005	Conducted Immunity
EN 61000-4-11	2004		Voltage Dips, Short Interruptions and Voltage Variations Immunity
EN 61000-3-2	2006		Harmonic Current Emissions
IEC 61000-3-3 <sup>1</sup>	2005		Voltage Fluctuations and Flicker

### **2006/95/EC Low-Voltage Directive**

<b>EN 61010-1</b>	<b>2001</b>		<b>Safety Requirements of electrical equipment for measurement, control and laboratory use. Part 1: General requirements</b>
-------------------	-------------	--	--

We, the undersigned, hereby declare that the equipment specified above conforms to the above directive(s).  
Manufacturer: Glas-Col

  
Signature

Steve Sterrett

President  
Position

Terre Haute, IN  
Place

10-15-12  
Date

**Glas-Col, LLC**  
711 Hulman Street., P.O. Box 2128,  
Terre Haute, IN 47802-0128  
Phone. (812) 235-6167,  
Fax. (812) 234-6975  
E-mail: [pinnacle@glascol.com](mailto:pinnacle@glascol.com)  
Web: [www.glascol.com](http://www.glascol.com)