



QUALITY CONTROL RANGES

MILLIPLEX[®] MAP Human Liver Injury Magnetic Bead Panel Kit Catalog # HLINJMAG-75K

Control Catalog # HLINJ-6075 Lot# HLI-104 and HLI-204

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
ARG1	Control I	467 – 969	pg/mL
	Control II	4020 – 8348	pg/mL
MDH1	Control I	458 – 951	pg/mL
	Control II	3843 – 7982	pg/mL
GSTa	Control I	80 – 167	pg/mL
	Control II	692 – 1437	pg/mL
5-NT	Control I	241 – 500	pg/mL
	Control II	2071 – 4302	pg/mL
SDH	Control I	3970 – 8246	pg/mL
	Control II	33212 – 68979	pg/mL

HLINJ-6075: HLI-104.204 (27 April 2021)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.