

QUALITY CONTROL RANGES
MILLIPLEX® MAP Human Cardiovascular Disease Magnetic Bead Panel 2
Kit Catalog# HCVD2MAG-67K
Control Catalog # HCVMG2-6067-2 Lot# HCV2-110.210



Analyte	QC Level	Expected Range	Units	Analyte	QC Level	Expected Range	Units
ADMTS13	Control 1	16 - 33	ng/mL	NGAL	Control 1	0.16 - 0.33	ng/mL
	Control 2	123 - 256	ng/mL		Control 2	1.3 - 2.7	ng/mL
D-dimer	Control 1	3.2 - 6.6	ng/mL	sVCAM-1	Control 1	1.7 - 3.6	ng/mL
	Control 2	26 - 54	ng/mL		Control 2	13 - 28	ng/mL
GDF-15	Control 1	0.008 - 0.02	ng/mL	SAA	Control 1	2.4 - 5.0	ng/mL
	Control 2	0.07 - 0.14	ng/mL		Control 2	22 - 45	ng/mL
Myoglobin	Control 1	0.30 - 0.63	ng/mL				
	Control 2	2.4 - 4.9	ng/mL				
sICAM-1	Control 1	1.2 - 2.4	ng/mL				
	Control 2	9.0 - 19	ng/mL				
MPO	Control 1	0.32 - 0.73	ng/mL				
	Control 2	2.6 - 5.3	ng/mL				
P-selectin	Control 1	2.8 - 5.8	ng/mL				
	Control 2	24 - 49	ng/mL				

HCVMG2-6067-2: HCV2-110.210 (27 August 2021)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.