

Specification – Certified Reference Material

Certipur® ICP Multi element standard solution IV

Accreditation:



Deutsche
Akkreditierungsstelle
D-RM-15185-01-00

Merck KGaA, Darmstadt, Germany is accredited by the German accreditation authority as registered reference material producer (D-RM-15185-01-00) in accordance with **ISO 17034**.

Producer: Merck KGaA, Frankfurter Str. 250, 64293 Darmstadt, Germany
Description of CRM: ICP Multi element standard solution IV
Ord. No.: 1.04498.0100
Expiry date: 3 years
Storage: +15°C to +25°C tightly closed in the original container
Matrix: HNO₃ 6.5%

Element	Specification	Associated uncertainty, $U=k \cdot u$ ($k=2$) as mass concentration	Traceable to NIST SRM®
Ag	900 - 1100 mg/l	±20 mg/l	SRM 3151
Al	900 - 1100 mg/l	±20 mg/l	SRM 3101a
B	900 - 1100 mg/l	±20 mg/l	SRM 3107
Ba	900 - 1100 mg/l	±20 mg/l	SRM 3104a
Bi	900 - 1100 mg/l	±20 mg/l	SRM 3106
Ca	900 - 1100 mg/l	±20 mg/l	SRM 3109a
Cd	900 - 1100 mg/l	±20 mg/l	SRM 3108
Co	900 - 1100 mg/l	±20 mg/l	SRM 3113
Cr	900 - 1100 mg/l	±20 mg/l	SRM 3112a
Cu	900 - 1100 mg/l	±20 mg/l	SRM 3114
Fe	900 - 1100 mg/l	±20 mg/l	SRM 3126a
Ga	900 - 1100 mg/l	±20 mg/l	SRM 3119a
In	900 - 1100 mg/l	±20 mg/l	SRM 3124a
K	900 - 1100 mg/l	±20 mg/l	SRM 3141a
Li	900 - 1100 mg/l	±20 mg/l	SRM 3129a
Mg	900 - 1100 mg/l	±20 mg/l	SRM 3131a
Mn	900 - 1100 mg/l	±20 mg/l	SRM 3132
Na	900 - 1100 mg/l	±20 mg/l	SRM 3152a
Ni	900 - 1100 mg/l	±20 mg/l	SRM 3136
Pb	900 - 1100 mg/l	±20 mg/l	SRM 3128
Sr	900 - 1100 mg/l	±20 mg/l	SRM 3153a
Tl	900 - 1100 mg/l	±20 mg/l	SRM 3158
Zn	900 - 1100 mg/l	±20 mg/l	SRM 3168a



Metrological traceability: This certified reference material has been measured applying high precision ICP-OES and is directly traceable to the corresponding **NIST SRM®** as mentioned on page 1.
NIST: National Institute of Standards and Technology, Gaithersburg, USA.

Measurement method: Inductively coupled plasma optical emission spectrometry ICP-OES.

Application and correct use: This certified reference material is intended for use as **calibration standard** for atomic spectrometry, spectrophotometry and other analytical techniques. Shake well before use and never pipet directly from the original container.

Associated uncertainty:
The associated uncertainty U_{CRM} reported with the certified values is calculated as combined expanded uncertainty $U_{\text{CRM}}=k \cdot u_{\text{CRM}}$ in accordance with GUM and EA-4/02, with $k=2$ as the coverage factor for a 95% coverage probability.

The combined uncertainty u_{CRM} is derived from combination of the squared uncertainty contributions:

$$u_{\text{CRM}} = \sqrt{u^2_{\text{Characterisation}} + u^2_{\text{Homogeneity}} + u^2_{\text{Stability}}}$$

$u_{\text{characterisation}}$: is the uncertainty in accordance with DIN EN ISO/IEC 17025 which includes the contributions of the primary reference material and the measuring system. The characterisation measurements have been conducted by our DAkkS accredited calibration laboratory.

$u_{\text{homogeneity}}$: is the between-bottle variation in accordance with ISO 17034. The assessment of homogeneity is performed by analysis of a representative number of systematically chosen sample units.

$u_{\text{stability}}$: is the uncertainty obtained from short-term and long-term stability in accordance with ISO 17034. The stability studies are the basis for the quantification of the expiry date of this elemental standard for the unopened bottle.

Detailed information is provided by the certificates and the certification report on our website.

The vibrant M, Supelco, Certipur and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates.
Detailed information on trademarks is available via publicly accessible resources.
© 2024 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

