



QUALITY CONTROL RANGES
MILLIPLEX[®] MAP Human MMP Magnetic Bead Panel 1
Kit Catalog # HMMP1MAG-55K

Control Catalog # HMMP1MG-6055-1
Lot# HMMPM-110 and HMMPM-210

<u>Hormone</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
HMMP-3	Control I	786 - 1633	pg/mL
	Control II	7531 - 15643	pg/mL
HMMP-12	Control I	671 - 1394	pg/mL
	Control II	5180 - 10759	pg/mL
HMMP-13	Control I	389 - 808	pg/mL
	Control II	3154 - 6551	pg/mL

Note: The Quality Control Ranges are generated with same day assay format using serum matrix provided in the kit. Quality Control values in culture media are not tested.

HMMP1MG-6055-1: HMMPM-110.210 (28 February 2024)
EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com
The M logo is a trademark of Merck KGaA, Darmstadt, Germany.
© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA