



QUALITY CONTROL RANGES

CANINE C-PEPTIDE ELISA CATALOG # EZCCP-47K AND EZCCP-47BK

STANDARD, LYOPHILIZED CATALOG#: E8047-K

STANDARD	C-PEPTIDE VALUE
Lot Number. 8047-6K	11.1 ng/ml

The exact value of lyophilized standard is shown above.

Reconstitute with **0.5 mL** distilled or deionized water.

QUALITY CONTROLS 1 AND 2, LYOPHILIZED CATALOG #: E6047-K

Reconstitute each vial with **0.5 mL** distilled or deionized water.

CONTROL	LOT NUMBER	EXPECTED RANGE	UNITS
QC 1	CCP-106	0.35 – 0.65	ng/ml
QC 2	CCP-206	1.5 – 3.0	ng/ml

E6047-K: CCP-106.206 (17 March 2022)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.