

## Specification – Certified Reference Material

### Certipur® ICP Multi element standard solution XVII

#### Accreditation:



Deutsche  
Akkreditierungsstelle  
D-RM-15185-01-00

Merck KGaA, Darmstadt, Germany is accredited by the German accreditation authority as registered reference material producer (D-RM-15185-01-00) in accordance with **ISO 17034**.

<b>Producer:</b>	Merck KGaA, Frankfurter Str. 250, 64293 Darmstadt, Germany
<b>Description of CRM:</b>	ICP Multi element standard solution XVII
<b>Ord. No.:</b>	1.04486.0100
<b>Expiry date:</b>	3 years
<b>Storage:</b>	+15°C to +25°C tightly closed in the original container
<b>Matrix:</b>	HCl 15%

Element	Specification	Associated uncertainty, $U=k \cdot u$ ( $k=2$ ) as mass concentration	Traceable to NIST SRM®
Hf	90 - 110 mg/l	±5 mg/l	SRM 3122
Ir	90 - 110 mg/l	±5 mg/l	Internal standard*
Sb	90 - 110 mg/l	±5 mg/l	SRM 3102a
Sn	90 - 110 mg/l	±5 mg/l	SRM 3161a
Ta	90 - 110 mg/l	±5 mg/l	SRM 3155
Ti	90 - 110 mg/l	±5 mg/l	SRM 3162a
Zr	90 - 110 mg/l	±5 mg/l	SRM 3169

\* SRM® NIST is not available



**Metrological traceability:**

This certified reference material has been measured applying high precision ICP-OES and is directly traceable to the corresponding **NIST SRM®** as mentioned on page 1.

*NIST: National Institute of Standards and Technology, Gaithersburg, USA.*

**Measurement method:**

Inductively coupled plasma optical emission spectrometry ICP-OES.

**Application and correct use:**

This certified reference material is intended for use as **calibration standard** for atomic absorption spectrometry, spectrophotometry and other analytical techniques. Shake well before use and never pipet directly from the original container.

**Associated uncertainty:**

The associated uncertainty  $U_{CRM}$  reported with the certified values is calculated as combined expanded uncertainty  $U_{CRM}=k \cdot u_{CRM}$  in accordance with GUM and EA-4/02, with  $k=2$  as the coverage factor for a 95% coverage probability.

The combined uncertainty  $u_{CRM}$  is derived from combination of the squared uncertainty contributions:

$$u_{CRM} = \sqrt{u_{Characterisation}^2 + u_{Homogeneity}^2 + u_{Stability}^2}$$

**Ucharacterisation:**

is the uncertainty in accordance with DIN EN ISO/IEC 17025 which includes the contributions of the primary reference material and the measuring system. The characterisation measurements have been conducted by our DAkkS accredited calibration laboratory.

**Uhomogeneity:**

is the between-bottle variation in accordance with ISO 17034. The assessment of homogeneity is performed by analysis of a representative number of systematically chosen sample units.

**Ustability:**

is the uncertainty obtained from short-term and long-term stability in accordance with ISO 17034. The stability studies are the basis for the quantification of the expiry date of this elemental standard for the unopened bottle.

**Detailed information is provided by the certificates and the certification report on our website.**

