3050 Spruce Street, St. Louis, MO 63103 USA
Tel: (800) 521-8956 (314) 771-5765 Fax: (800) 325-5052 (314) 771-5757
email: techservice@sial.com sigma-aldrich.com

# **Product Information**

Cannabinol solution, 1.0 mg/mL±5% in methanol drug standard

Catalog Number **C6520** Lot Number 058K8706 Storage Temperature 2–8 °C

CAS RN 512-35-7 Synonyms: 6,6,9-Trimethyl-3-pentyl-6H-dibenzo[b,d]pyran-1-ol; CBN

#### **Product Description**

Molecular Formula: C<sub>21</sub>H<sub>26</sub>O<sub>2</sub> Molecular Weight: 310.43

DEA Class: Exempt preparation of Class I

Concentration: 0.98 mg of Cannabinol/mL of GC grade methanol, concentration verified by UV.

Expiration Date: April 2012

#### **Precautions and Disclaimer**


This product is for R&D use only, not for drug, household, or other uses. Please consult the Material Safety Data Sheet for information regarding hazards and safe handling practices.

## Storage/Stability


Store the product at  $2-8\,^{\circ}$ C. If stored at  $2-8\,^{\circ}$ C in an airtight container, protected from light, decomposition of the product is less than 1% in 48 months. After opening, concentration may change due to loss of solvent.

CWM,MAM 04/08-1

#### 70 eV Electron Ionization Mass Spectrum



Hewlett-Packard 5890 Series II Gas Chromatograph Hewlett-Packard 5971 Mass Selective Detector Hewlett-Packard G1034C MS ChemStation (HP-DOS series)



Nicolet Magna IR Spectrometer Series II

## **Ultraviolet Spectrum**

Peak 285 nm

Absorbance 0.3108

**Solvent** GC grade methanol


**Dilution** 1:199 (v/v)


## **Capillary Gas Chromatography**

Column Supelco SPB-1 (Catalog Number 24162) I.D. 0.20 mm Length 15 m Film Thickness 0.20  $\mu$ m Oven Temperature 50 °C (2 minutes), then 15 °C/minute to 300 °C, hold 11 minutes

**Injector Temperature** 250 °C, Split **Detector** Hewlett-Packard 5971

Mass Selective Detector (70 eV, 40-500 m/z)





Bio-Tek Uvikon XL