


**SIGMA QUALITY CONTROL TEST PROCEDURE****Product Information****Enzymatic Assay of ADENOSINE 5'-TRIPHOSPHATASE  
(EC 3.6.1.3)  
From Dog and Rabbit Kidney****PRINCIPLE:**

Abbreviations used:

ATPase = Adenosine 5'-Triphosphatase

ATP = Adenosine 5'-Triphosphate

ADP = Adenosine 5'-Diphosphate

P<sub>i</sub> = Inorganic Phosphate

**CONDITIONS:** T = 37°C, pH 7.4, A<sub>660nm</sub>, Light path = 1 cm

**METHOD:** Colorimetric

**REAGENTS:**

- A. 50 mM Tris HCl Buffer, pH 7.4 at 37°C  
(Prepare 100 ml in deionized water using Trizma Base, Sigma Prod. No. T-1503. Adjust to pH 7.4 at 37°C with 1 M HCl.)
- B. 15 mM Ouabain Solution  
(Prepare 10 ml in deionized water using Ouabain, Octahydrate Sigma Prod. No. O-3125.)
- C. 600 mM Potassium Chloride Solution (KCl)  
(Prepare 10 ml in deionized water using Potassium Chloride, Sigma Prod. No. P-4504.)
- D. 2.5 M Sodium Chloride Solution (NaCl)  
(Prepare 10 ml in deionized water using Sodium Chloride, Sigma Prod. No. S-9625.)
- E. 90 mM Magnesium Chloride Solution (MgCl<sub>2</sub>)  
(Prepare 10 ml in deionized water using Magnesium Chloride Hexahydrate, Sigma Prod. No. M-0250.)

**Enzymatic Assay of ADENOSINE 5'-TRIPHOSPHATASE  
(EC 3.6.1.3)  
From Dog and Rabbit Kidney**

**REAGENTS:** (continued)

- F. 80 mM Adenosine 5'-Triphosphate (ATP)  
(Prepare 10 ml in deionized water using Adenosine 5'-Triphosphate, Tris, Sigma Prod. No. A-9062. Adjust to pH 7.4 at 37°C with 1 M Tris.)
- G. 10% (w/v) Ammonium Molybdate Solution  
(Prepare 25 ml in 10 N H<sub>2</sub>SO<sub>4</sub> using Molybdic Acid, Ammonia Tetrahydrate Salt, Sigma Prod. No. M-0878.)
- H. Taussky-Shorr Color Reagent (TSCR)  
(Prepare by adding 10 ml of Reagent G to 70 ml of deionized water. Add 5 g of Ferrous Sulfate, Heptahydrate, Sigma Prod. No. F-0131. Bring the volume to 100 ml with deionized water.)
- I. 20% (w/v) Trichloroacetic Acid (TCA)  
(Prepare 100 ml in deionized water using Trichloroacetic Acid, 6.1 N, approximately 100% (w/v), Sigma Stock No. 490-10.)
- J. Phosphorus Standard (P Std)  
(Use Phosphorus Standard Solution, Sigma Prod. No. 661-9. The concentration is 20 µg/ml, 0.645 µmoles/ml.)
- K. Adenosine 5'-Triphosphatase Enzyme Solution  
(Immediately before use, prepare a solution in cold deionized water containing 4 - 5 units<sup>1</sup>/ml.)

**PROCEDURE:**

Step 1:

Pipette (in milliliters) the following reagents into suitable containers:

	<u>Test 1</u>	<u>Blank 1</u>	<u>Test 2</u>	<u>Blank 2</u>
Reagent A (Buffer)	0.90	0.90	0.90	0.90
Reagent B (Ouabain)	-----	-----	0.10	0.10
Reagent C (KCl)	0.05	0.05	0.05	0.05
Reagent D (NaCl)	0.10	0.10	0.10	0.10
Reagent E (MgCl <sub>2</sub> )	0.05	0.05	0.05	0.05
Reagent K (Enzyme)	0.10	-----	0.10	-----
Deionized Water	0.20	0.20	0.10	0.10

Mix and equilibrate for 5 minutes at 37°C. Then add:

Reagent F (ATP)	0.10	0.10	0.10	0.10
-----------------	------	------	------	------

**Enzymatic Assay of ADENOSINE 5'-TRIPHOSPHATASE  
(EC 3.6.1.3)  
From Dog and Rabbit Kidney**

**PROCEDURE:** (continued)

Mix and incubate at 37°C for exactly 5 minutes. Then add:

	<u>Test 1</u>	<u>Blank 1</u>	<u>Test 2</u>	<u>Blank 2</u>
Reagent I (TCA)	1.50	1.50	1.50	1.50

Immediately mix by inversion. Then add:

Reagent K (Enzyme Solution)	-----	0.10	-----	0.10
--------------------------------	-------	------	-------	------

Mix by inversion and then centrifuge in a clinical centrifuge for 3 minutes to clarify.

Step 2:

Pipette (in milliliters) the following into suitable tubes:

Reagent H (TSCR)	2.00	2.00	2.00	2.00
Test Supernatant	1.00	-----	1.00	-----
Blank Supernatant	-----	1.00	-----	1.00
Deionized Water	1.00	1.00	1.00	1.00

Mix and incubate at 25°C for 10 minutes. Read the  $A_{660\text{nm}}$  for both Tests and Blanks.

Standard Curve:

Prepare a standard curve by pipetting the following into suitable tubes (milliliters).

	<u>Std 1</u>	<u>Std 2</u>	<u>Std 3</u>	<u>Blank</u>
Reagent H (TSCR)	2.00	2.00	2.00	2.00
Reagent J (P Std)	0.25	0.50	1.00	-----
Reagent I (TCA)	0.50	0.50	0.50	0.50
Deionized Water	1.25	1.00	0.50	1.50

Mix and incubate at 25°C for 10 minutes. Read the  $A_{660\text{nm}}$  for each standard using a suitable spectrophotometer.

**Enzymatic Assay of ADENOSINE 5'-TRIPHOSPHATASE  
(EC 3.6.1.3)  
From Dog and Rabbit Kidney**

**CALCULATIONS:**

Standard Curve:

$$\Delta A_{660\text{nm}} \text{ Standard} = A_{660\text{nm}} \text{ Standard} - A_{660\text{nm}} \text{ Standard Blank}$$

Prepare a standard curve by plotting  $\Delta A_{660\text{nm}}$  Standard vs Phosphate concentration.

Sample Determination:

$$\Delta A_{660\text{nm}} \text{ Test}_1 = A_{660\text{nm}} \text{ Test}_1 - A_{660\text{nm}} \text{ Blank}_1$$

$$\Delta A_{660\text{nm}} \text{ Test}_2 = A_{660\text{nm}} \text{ Test}_2 - A_{660\text{nm}} \text{ Blank}_2$$

Determine the micromoles of Phosphate liberated for each Test using the standard curve.

$$\text{Units/ml enzyme} = \frac{(\mu\text{moles of Phosphate released})(3.0)(df)}{(5)(0.1)(1.0)}$$

1.0 = Aliquot (in milliliter) of Test Supernatant used in Step 2

3 = Total volume (in milliliters) of assay (Step 1)

5 = Time of assay (in minutes) as per the Unit Definition

0.1 = Volume (in milliliter) of enzyme used in Step 1

$$\text{Units/mg solid} = \frac{\text{units/ml enzyme}}{\text{mg solid/ml enzyme}}$$

$$\text{Units/mg protein} = \frac{\text{units/ml enzyme}}{\text{mg protein/ml enzyme}}$$

Test 1 = ATPase, activated (Na, K, Mg)

Test 2 = ATPase, not Ouabain sensitive

Ouabain sensitive = Test 1 - Test 2

**UNIT DEFINITION:**

One unit will liberate 1.0  $\mu$ mole of inorganic phosphorus from ATP per minute at pH 7.4 at 37°C in the presence of  $\text{Na}^+$ ,  $\text{K}^+$ , and  $\text{Mg}^{++}$ .

**Enzymatic Assay of ADENOSINE 5'-TRIPHOSPHATASE  
(EC 3.6.1.3)  
From Dog and Rabbit Kidney**

**FINAL ASSAY CONCENTRATION:**

In a 1.50 ml reaction mix, the final concentrations are 30 mM Tris, 3 mM magnesium chloride, 5.3 mM adenosine 5'-triphosphate, 167 mM sodium chloride, 20 mM potassium chloride, 1 mM ouabain (when present), and 0.1 - 0.2 unit ATPase.

**REFERENCES:**

Bonting, S.L., Simon, K.A., and Hawkins, N.M. (1961) *Archives of Biochemistry and Biophysics* **95**, 416-423

Taussky, H.H. and Shorr, E. (1953) *Journal of Biological Chemistry* **202**, 675-685

**NOTES:**

1. These units are the fully active (Test 1) units.
2. This assay is based on the cited references.
3. Where Sigma Product or Stock numbers are specified, equivalent reagents may be substituted.

**Sigma warrants that the above procedure information is currently utilized at Sigma and that Sigma products conform to the information in Sigma publications. Purchaser must determine the suitability of the information and products for its particular use. Upon purchase of Sigma products, see reverse side of invoice or packing slip for additional terms and conditions of sale.**