

**Technical Bulletin** 

# Acid Phosphatase Fluorometric Assay Kit

#### Catalogue number MAK525

# **Product Description**

Acid Phosphatase (ACP) is an enzyme which catalyzes the cleavage of phosphate groups from other molecules during digestion. Acid phosphatase can be found in lysosomes and become active after fusing with endosomes, acidifying the pH and, thus, creating an optimal environment for ACP. ACP can also be found in bone, spleen, liver, kidney and blood.

The Acid Phosphatase Fluorometric Assay Kit is a non-radioactive, fluorometric ACP assay is based on the cleavage of a synthetic substrate. The product methylumbelliferone (MUB) becomes intensely fluorescent after addition of the stop reagent. The increase in fluorescence ( $\lambda_{Ex}=360~\text{nm}/\lambda_{Em}=450~\text{nm}$ ) after addition of the stop reagent is directly proportional to the enzyme activity.

The linear detection range of the kit is (20  $\mu$ L sample): 0.008 to 10 U/L for a 30-minute reaction. The kit is used to determine Acid Phosphatase activity in biological samples, such as plasma, serum, cell lysate, tissue samples.

## Components

The kit is sufficient for 100 fluorometric assays in 96-well plates.

•	Assay Buffer	12 mL
	Catalogue Number MAK525A MUP Substrate	120 µL
	Catalogue Number MAK525B	120 μΕ
•	Stop Reagent	12 mL
	Catalogue Number MAK525C	
•	MUB Standard	120 µL
	Catalogue Number MAK525D	

# Reagents and Equipment Required but Not Provided

- Pipetting devices and accessories (for example, multichannel pipettor)
- Fluorescent multiwell plate reader
- Black flat-bottom 96-well or 384-well plates. Cell culture or tissue culture treated plates are not recommended.
- 1.5 mL microcentrifuge tubes
- Dounce tissue grinder set (Catalogue Number D9063 or equivalent).

## Precautions and Disclaimer

For Research Use Only. Not for use in diagnostic procedures. Please consult the Safety Data Sheet for information regarding hazards and safe handling practices.

# Storage/Stability

1

The kit is shipped at room temperature. Store components at -20 °C.

## **Preparation Instructions**

Briefly centrifuge small vials prior to opening. Equilibrate all components to room temperature prior to use.

**Note**: This assay is based on a kinetic reaction. To ensure identical incubation time, addition of Substrate and Stop Reagent to samples should be quick, and mixing should be brief but thorough. Use of a multichannel pipettor is recommended.



## Sample Preparation

Serum and plasma should be diluted 2-5 fold.

#### Tissue:

- 1. Prior to dissection, rinse tissue in Tris buffered saline (pH 7.4) to remove blood.
- 2. Homogenize tissue (50 mg) in  $\sim$  200  $\mu$ L 50 mM Tris buffer (pH 7.5).
- 3. Centrifuge at 14,000 x g for 10 minutes at 4 °C.
- 4. Remove supernatant for assay.

#### Cell Lysate:

1. Collect cells by centrifugation at 2,000 x g for 5 minutes at 4 °C.

**Note:** For adherent cells, do not harvest cells using proteolytic enzymes; rather use a rubber policeman.

- 2. Homogenize or sonicate cells in an appropriate volume of cold 50 mM Tris buffer (pH 7.5), approximately one million cells per mL.
- 3. Centrifuge at 14,000 x g for 10 minutes at 4 °C.
- 4. Remove supernatant for assay.

**Note:** All samples can be stored at - 80 to 20 °C for at least one month.

## Procedure

All Samples and Standards should be run in duplicate. **Note:** This procedure is written for 96-well plates.

## Standard Curve Preparation

- 1. Prepare 100  $\mu$ M standard by mixing 25  $\mu$ L of Standard (MUB) with 475  $\mu$ L purified water.
- 2. Dilute the 100 μM standard according to Table 1.

Table 1

. 45.6 =						
Well No.	Standard (µL)	Purified water (µL)	MUB Conc. (µM)			
1	100	0	100			
2	60	40	60			
3	30	70	30			
4	0	100	0			

## **Reaction Preparation**

- 1. Transfer 20 µL of each sample into separate wells.
- 2. Transfer 20 μL of each standard into wells of a black flat-bottom 96-well plate.

## Working Reagent Preparation

Mix enough reagents for the number of assays to be performed. For each Standard and Sample well, prepare 80  $\mu L$  of Working Reagent according to Table 2.

#### Table 2.

Preparation of Working Reagent

Reagent	Working Reagent
Assay Buffer	85 µL
MUP Substrate	1 μL

- 1. Add 80  $\mu L$  of Working Reagent to all Standard and Sample wells.
- 2. Tap plate briefly to mix.

## Measurement

- Incubate at 25 °C or desired temperature for 30 minutes.
- 2. Add 50  $\mu$ L of Stop Reagent to each well. Tap plate briefly to mix.
- 3. Read fluorescence (F) at  $\lambda_{Ex} = 360 \text{ nm}/\lambda_{Em} = 450 \text{ nm}$ .

#### Results

- 1. Subtract blank fluorescence (F) (standard #4) from the standard fluorescence values and plot the  $\Delta F$  against standard concentrations.
- 2. Determine the slope.
- 3. Use following equation to calculate Acid Phosphatase Activity.

Acid Phosphatase Activity (U/L) =

$$\frac{F_{Sample}\text{-} F_{Blank}}{\text{Time x Slope}} \times \mathsf{DF}$$

#### Where:

 $F_{SAMPLE}$  = RFU value for each sample

 $F_{BLANK} = RFU$  value for water (standard 4)

Slope = slope of the linear regression fit of the standard points


Time = reaction time (30 min)

DF = Dilution Factor.


Unit definition: 1 Unit (U) of ACP will catalyze the conversion of 1  $\mu$ mole of 4-methylumbelliferyl phosphate to 4-methylumbelliferone and phosphate per min at 25 °C and pH 5.3.

**Note**: If sample ACP activity exceeds 10 U/L, either use a shorter reaction time or dilute samples in water and repeat the assay. For samples with ACP activity < 0.1 U/L, the incubation time can be extended up to 60 minutes for greater sensitivity.

**Figure 1.** Typical Titration curve



**Figure 2.** Typical Standard Curve



#### Notice

We provide information and advice to our customers on application technologies and regulatory matters to the best of our knowledge and ability, but without obligation or liability. Existing laws and regulations are to be observed in all cases by our customers. This also applies in respect to any rights of third parties. Our information and advice do not relieve our customers of their own responsibility for checking the suitability of our products for the envisaged purpose.

The information in this document is subject to change without notice and should not be construed as a commitment by the manufacturing or selling entity, or an affiliate. We assume no responsibility for any errors that may appear in this document.

#### **Contact Information**

For the location of the office nearest you, go to SigmaAldrich.com/offices.

#### **Technical Service**

Visit the tech service page on our web site at <a href="SigmaAldrich.com/techservice">SigmaAldrich.com/techservice</a>.

#### Standard Warranty

The applicable warranty for the products listed in this publication may be found at SigmaAldrich.com/terms.

The life science business of Merck operates as MilliporeSigma in the U.S. and Canada.

