



QUALITY CONTROL RANGES

HUMAN GHRELIN (ACTIVE) ELISA KIT CATALOG #: EZGRA-88K

HUMAN GHRELIN ACTIVE STANDARD, LYOPHILIZED CATALOG #: E8088-K

STANDARD	HUMAN GHRELIN VALUE
Lot Number: EGA-17K	1880 pg/mL

The exact value of lyophilized standard is shown above.

Reconstitute with **2.0mL** distilled or deionized water.

QUALITY CONTROLS 1 AND 2, LYOPHILIZED CATALOG #: E6088-K

Reconstitute each vial with **0.5 mL** distilled or deionized water.

CONTROL	LOT NUMBER	EXPECTED RANGE	UNITS
QC 1	AGR-117	93 – 194	pg/mL
QC 2	AGR-217	455 – 945	pg/mL

E6088-K: AGR-117, 217 (13 September 2023)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA