



QUALITY CONTROL RANGES

Human C-Peptide ELISA

Catalog #: EZHCP-20K

Control Catalog #: E6020-K

CONTROL	LOT NUMBER	EXPECTED RANGE	UNITS
QC 1	HCP-115	0.38 - 0.80	ng/mL
QC 2	HCP-215	3.4 - 7.0	ng/mL

E6020-K: HCP-115.215 (01 April 2020)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.