3050 Spruce Street, St. Louis, MO 63103 USA
Tel: (800) 521-8956 (314) 771-5765 Fax: (800) 325-5052 (314) 771-5757
email: techservice@sial.com sigma-aldrich.com

# **Product Information**

(–)-Cotinine solution 1.0 mg/mL±5% in methanol drug standard

Catalog Number **C0430**Lot Number 061M8703
Storage Temperature 2–8 °C

CAS RN 486-56-6

Synonyms: (*S*)-1-Methyl-5-(3-pyridyl)-2-pyrrolidinone, *S*(-)-1-Methyl-5-(3-pyridyl)-2-pyrrolidinone

# **Product Description**

Molecular Formula: C<sub>10</sub>H<sub>12</sub>N<sub>2</sub>O Molecular Weight: 176.22

DEA Class: Exempt preparation of a non-controlled

substance

Concentration: 1.00 mg of (–)-cotinine/mL of GC grade methanol, concentration verified by UV.

Expiration Date: May 2014

#### **Precautions and Disclaimer**

This product is for R&D use only, not for drug, household, or other uses. Please consult the Material Safety Data Sheet for information regarding hazards and safe handling practices.


## Storage/Stability

Store the product at 2-8 °C.


If stored at 2–8 °C in an airtight container, protected from light, decomposition of the product is less than 1% in 12 months. After opening, concentration may change due to loss of solvent.

VY,CMW,MAM 01/16-1

#### 70 eV Electron Ionization Mass Spectrum



Hewlett-Packard 5890 Series II Gas Chromatograph Hewlett-Packard 5971 Mass Selective Detector Hewlett-Packard G1034C MS ChemStation (HP-DOS series)



## Nicolet 380 FT-IR Thermo Smart Orbit

# **Ultraviolet Spectrum**

**Peak** 261.0 nm


Absorbance 0.7406


**Solvent** GC grade methanol

**Dilution** 1:24 (v/v)

## **Capillary Gas Chromatography**

Column Supelco SPB-1 (Catalog Number 24162)
I.D. 0.20 mm Length 15 m Film Thickness 0.20 μm
Oven Temperature 50 °C (2 minutes), then
15 °C/minute to 300 °C, hold 11 minutes
Injector Temperature 250 °C, Split
Detector Hewlett-Packard 5971
Mass Selective Detector (70 eV, 40-500 m/z)





Beckman-Coulter DU 800