



Quality Control Ranges
MILLIPLEX[®] MAP Human Neuroscience
Magnetic Bead Panel 1
Kit Cat# HNS1MAG-95K
Control Cat# HNS1-6095-1
Lot# HNS1-104 and HNS1-204

<u>Analyte</u>	<u>Control Level</u>	<u>Range</u>	<u>Units</u>
GFAP	Control 1	161 – 334	pg/mL
	Control 2	838 – 1741	pg/mL
UCHL1	Control 1	422 – 877	pg/mL
	Control 2	2165 – 4496	pg/mL
DJ1	Control 1	1375 – 2855	pg/mL
	Control 2	7414 – 15398	pg/mL
NSE	Control 1	296 – 614	pg/mL
	Control 2	1532 – 3182	pg/mL
α -Synuclein	Control 1	4157 – 8634	pg/mL
	Control 2	22104 – 45908	pg/mL
TGM2	Control 1	1123 – 2332	pg/mL
	Control 2	6126 – 12724	pg/mL

HNS1-6095-1: HNS1-104.204 (17 June 2021)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.