

Quality Control Ranges

MILLIPLEX® PORCINE ACUTE PHASE PROTEIN

Catalog Number PAPP1-160K

QC Catalog Number PAPP1-6100 / Lot Number 6160-102 and 6160-202

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
Haptoglobin	Control I	9 - 18	ng/mL
	Control II	36 - 74	ng/mL
CRP	Control I	5 - 11	ng/mL
	Control II	21 - 44	ng/mL

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.

The vibrant M, Millipore, Millipore and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates.

© 2025 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

Cat#: PAPP1-6100, Lot#: 6160-102 and 6160-202 (15DEC2025)

