



**Quality Control Ranges
Rat Bone Magnetic Bead Panel 1
Kit Cat# RBN1MAG-31K**

**Control Cat# RBN-6031
Lot# RBN31-109 and RBN31-209**

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
ACTH	Control 1	295 – 613	pg/ml
	Control 2	1150 – 2388	pg/ml
OPG	Control 1	318 – 660	pg/ml
	Control 2	1300 – 2701	pg/ml
Insulin	Control 1	1041 – 2161	pg/ml
	Control 2	4134 – 8585	pg/ml
Leptin	Control 1	242 – 504	pg/ml
	Control 2	1017 – 2113	pg/ml
PTH	Control 1	11 – 24	pg/ml
	Control 2	45 – 94	pg/ml
DKK1	Control 1	364 – 755	pg/ml
	Control 2	1500 – 3115	pg/ml
SOST	Control 1	146 – 304	pg/ml
	Control 2	614 – 1275	pg/ml
FGF-23	Control 1	714 – 1482	pg/ml
	Control 2	3064 – 6363	pg/ml

RBN-6031: RBN31-109.209 (28 May 2021)

EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com

The M logo is a trademark of Merck KGaA, Darmstadt, Germany.

© 2013 EMD Millipore Corporation, Billerica, MA 01821 USA.