



Product Information

SIGMA QUALITY CONTROL TEST PROCEDURE Enzymatic Assay of β -GLUCOSIDASE (EC 3.2.1.21)

PRINCIPLE:

β -D-Glucoside + H₂O $\xrightarrow{\beta\text{-Glucosidase}}$ D-Glucose + an Alcohol

CONDITIONS: T = 37°C, pH = 5.0, A_{540nm}, Light path = 1 cm

METHOD: Colorimetric¹

REAGENTS:

- A. 100 mM Sodium Acetate Buffer, pH 5.0 at 37°C
(Prepare 200 ml in deionized water using Sodium Acetate, Trihydrate, Sigma Prod. No. S-8625. Adjust to pH 5.0 at 37°C with 1 M HCl.)
- B. 1% (w/v) Salicin Substrate Solution (Salicin)
(Prepare 50 ml in Reagent A using Salicin, Sigma Prod. No. S-0625.)
- C. β -Glucosidase Enzyme Solution
(Immediately before use, prepare a solution containing 1.2 - 2.4 units/ml of β -Glucosidase in cold deionized water.)
- D. 16 mM Copper Sulfate, 1300 mM Sodium Sulfate,
226 mM Sodium Carbonate, 190 mM Sodium Bicarbonate, and 43 mM Sodium Potassium Tartrate Solution (Copper Soln)
(Prepare 1 liter in deionized water using Cupric Sulfate Pentahydrate, Sigma Prod. No. C-7631, Sodium Bicarbonate, Sigma Prod. No. S-8875, Sodium Sulfate, Anhydrous, Sigma Prod. No. S-9627, Sodium Carbonate, Anhydrous, Sigma Prod. No. S-2127, and Sodium Potassium Tartrate Tetrahydrate, Sigma Prod. No. S-2377.²)
- E. 40 mM Molybdic Acid, 19 mM Arsenic Acid, and
756 mM Sulfuric Acid Solution (Ars-Mol Soln)
(Prepare 1 liter in deionized water using Molybdic Acid, Ammonium Salt Tetrahydrate, Sigma Prod. No. M-0878, Arsenic Acid, Sodium Salt, Sigma Prod. No. A-6756, and Sulfuric Acid, Sigma Prod. No. S-1526.³)

**Enzymatic Assay of β -GLUCOSIDASE
(EC 3.2.1.21)**

REAGENTS: (continued)

- F. Glucose Standard Solution (Glucose)
(Use Glucose Standard Solution, Sigma Stock No. 635-100.)

PROCEDURE:

Pipette (in milliliters) the following reagents into suitable test tubes:

	<u>Test</u>	<u>Blank</u>
Reagent B (Salicin)	4.00	4.00

Equilibrate to 37°C. Then add:

Deionized Water	-----	1.00
Reagent C (Enzyme)	1.00	-----

Mix by inversion and incubate at 37°C for exactly 10 minutes.

Immediately transfer 1 ml of reaction mixture into a suitable container containing 1 ml of Reagent D as indicated below and proceed with Somogyi's method² of assaying reducing sugars.

Pipette (in milliliters) the following reagents into suitable containers:

	<u>Test</u>	Test <u>Blank</u>	<u>Std 1</u>	<u>Std 2</u>	<u>Std 3</u>	<u>Std 4</u>	<u>Std 5</u>	Std <u>Blank</u>
Test Solution	1.00	---	---	---	---	---	---	---
Test Blank Solution	---	1.00	---	---	---	---	---	---
Deionized Water	---	---	0.97	0.95	0.93	0.90	0.80	1.00
Reagent F (Glucose)	---	---	0.03	0.05	0.07	0.10	0.20	---
Reagent D (Copper Soln)	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00

Immediately mix by inversion. Place a marble over the top of the tube and transfer the tubes to a boiling water bath. Incubate for 10 minutes. Remove from the boiling water bath and allow to cool to room temperature. Then add:

Reagent E (Ars-Mol Soln)	1.00	1.00	1.00	1.00	1.00	1.00	1.00	1.00
--------------------------	------	------	------	------	------	------	------	------

**Enzymatic Assay of β -GLUCOSIDASE
(EC 3.2.1.21)**

PROCEDURE: (continued)

Shake or vortex until foaming stops and any precipitate present is dissolved. Then add:

	<u>Test</u>	<u>Test Blank</u>	<u>Std 1</u>	<u>Std 2</u>	<u>Std 3</u>	<u>Std 4</u>	<u>Std 5</u>	<u>Std Blank</u>
Deionized Water	10.00	10.00	10.00	10.00	10.00	10.00	10.00	10.00

Mix and transfer to suitable cuvettes. Obtain the A_{540nm} for Test, Blank and Standards, using a suitable spectrophotometer.

CALCULATIONS:

Standard Curve:

$$\Delta A_{540nm} \text{ Std} = A_{540nm} \text{ Std} - A_{540nm} \text{ Blank}$$

Prepare a standard curve by plotting the ΔA_{540nm} of the Standard versus the μ moles of glucose liberated.

Sample Determination:

$$\Delta A_{540nm} \text{ Sample} = A_{540nm} \text{ Test} - A_{540nm} \text{ Test Blank}$$

Determine the μ moles of glucose liberated using the Standard curve.

$$\text{Units/mg enzyme} = \frac{(\mu\text{moles of glucose liberated}) (5)}{(10) (1) (\text{mg enzyme/RM})}$$

5 = Volume of Reaction Mix

10 = Time of assay (in minutes) as per Unit Definition

1 = Volume of enzyme assay used in Colorimetric Determination

UNIT DEFINITION:

One unit will liberate 1.0 μ mole of glucose from salicin per minute at pH 5.0 at 37°C.

Enzymatic Assay of β -GLUCOSIDASE (EC 3.2.1.21)

FINAL ASSAY CONCENTRATIONS:

In a 5.00 ml reaction mix, the final concentrations are 80 mM sodium acetate, 0.8% (w/v) salicin and 1.2 - 2.4 units β -glucosidase.

REFERENCES:

- Somogyi M., (1952) *J. Biol. Chem.* **195**, 19-23.
Somogyi M., (1945) *J. Biol. Chem.* **160**, 61-68.
Nelson N., (1944) *J. Biol. Chem.* **153**, 375-380.

NOTES:

1. The method of assaying for the presence of reducing sugars, described here, is that of Somogyi.
2. Sodium Sulfate, Sodium Carbonate, and Sodium Potassium Tartrate are dissolved in approximately 500 ml of deionized water. Cupric Sulfate is dissolved in approximately 100 ml of deionized water and slowly added to the above solution to avoid precipitation. Sodium Bicarbonate is dissolved first in deionized water and then added to the above solution. Dilute the solution to 1 liter. If a precipitate forms, it should be removed by filtration prior to use. Store in an amber bottle and avoid exposure to direct sunlight. Store at room temperature.
3. Molybdc Acid is dissolved in approximately 300 ml of deionized water. Add Sulfuric Acid slowly. Caution, this is an exothermic reaction! A solution of arsenic acid is dissolved in approximately 300 ml of deionized water and is added to the above solution. The solution is diluted to a total volume of 1 liter and incubated at 37°C for 48 - 72 hours. If a precipitate forms, it should be removed by filtration prior to use. Store in an amber bottle and avoid exposure to direct sunlight. The solution expires six months after preparation. Store at room temperature in an exhaust hood.

**Enzymatic Assay of β -GLUCOSIDASE
(EC 3.2.1.21)**

NOTES: (continued)

4. All products and stock numbers, unless otherwise indicated, are Sigma product and stock numbers.

Sigma warrants that the above procedure information is currently utilized at Sigma and that Sigma products conform to the information in Sigma publications. Purchaser must determine the suitability of the information and products for its particular use. Upon purchase of Sigma products, see reverse side of invoice or packing slip for additional terms and conditions of sale.